



Designation: D4367 – 22

Standard Test Method for Benzene in Hydrocarbon Solvents by Gas Chromatography¹

This standard is issued under the fixed designation D4367; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the determination by gas chromatography of benzene at levels from 0.01 to 1 volume % in hydrocarbon solvents.

NOTE 1—For benzene levels lower than 0.01 volume %, use Test Method [D6229](#).

1.2 The values stated in SI units are to be regarded as the standard. The values given in parentheses are for information only.

1.3 For purposes of determining conformance of an observed or a calculated value using this test method to relevant specifications, test result(s) shall be rounded off “to the nearest unit” in the last right-hand digit used in expressing the specification limit, in accordance with the rounding-off method of Practice [E29](#).

1.4 For hazard information and guidance, see the supplier’s Material Safety Data Sheet. For specific hazard statements, see Section 7.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.6 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:²

¹ This test method is under the jurisdiction of ASTM Committee D01 on Paint and Related Coatings, Materials, and Applications and is the direct responsibility of Subcommittee D01.35 on Solvents, Plasticizers, and Chemical Intermediates.

Current edition approved Jan. 1, 2022. Published January 2022. Originally approved in 1984. Last previous edition approved in 2012 as D4367 – 02 (2012) which was withdrawn January 2021 and reinstated in January 2022. DOI: 10.1520/D4367-22.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard’s Document Summary page on the ASTM website.

[D6229 Test Method for Trace Benzene in Hydrocarbon Solvents by Capillary Gas Chromatography](#)
[E29 Practice for Using Significant Digits in Test Data to Determine Conformance with Specifications](#)
[E300 Practice for Sampling Industrial Chemicals](#)

3. Summary of Test Method

3.1 An internal standard, methyl ethyl ketone (MEK), is added to the material and then introduced into a gas chromatograph equipped with two columns connected in series. The specimen passes first through a column packed with the nonpolar phase, methyl silicone, which separates the components by boiling point. After octane has eluted, the flow through the nonpolar column is reversed, flushing out the components heavier than octane. The octane and lighter components then pass through a column with the highly polar phase, 1,2,3-tris(2-cyanoethoxy)propane, that separates the aromatic and nonaromatic compounds. The eluted components are detected by a conventional detector and recorded on a strip chart. The peak areas are measured and the concentration of each component is calculated by reference to the internal standard.

4. Significance and Use

4.1 Benzene is classed as a toxic and carcinogenic material. A knowledge of the concentration of this compound may be an aid in evaluating the possible health hazards to persons handling and using hydrocarbon solvents, but this test method is not intended to evaluate such hazards.

5. Apparatus

5.1 *Chromatograph*—Any gas chromatographic instrument that has a backflush system and flame ionization detector and that can be operated at the conditions given in [Table 1](#). The detector-recorder combination must produce a 4-mm deflection for a 1- μ L specimen containing 0.05 volume % MEK when operated at maximum sensitivity.

5.2 *Columns*, one 0.8-m (2.5-ft) length of 3.2-mm ($\frac{1}{8}$ -in.) outside diameter stainless steel tubing and one 4.6-m (15-ft) length of 3.2-mm ($\frac{1}{8}$ -in.) outside diameter stainless steel tubing.

5.3 *Recorder, Strip Chart*—Potentiometer with a full-scale deflection of 1 mV, a full-scale response time of 2 s or less, and a maximum noise level of ± 0.3 % of full scale.

TABLE 1 Instrument Conditions Found Satisfactory for Measuring Low Concentrations of Benzene in Hydrocarbon Solvents (Note 2)

Detector	flame ionization
Columns	two, stainless steel
Length, m	(A) 0.8; (B) 4.6
Outside diameter, mm	3.2
Stationary phases	(A) methyl silicone, 10 weight % (B) TCEP, 25 weight %
Support	(A) acid-washed calcined diatomite, 60 to 80-mesh (B) acid-washed pink diatomaceous earth, 80 to 100-mesh
Reference column	any column or restriction may be used
Temperature, °C	
Injection port	150
Column, isothermal	100
Detector block	150
Carrier gas	helium
Flow rate, mL/min	approximately 30
Recorder range, mV	0 to 1
Chart speed, mm/min	10
Specimen size, µL	1.0
Time to backflush, min	approximately 2
Total cycle time, min	approximately 30

5.4 *Microsyringe*, 5-µL capacity.

5.5 *Pipets*, measuring 1 and 2 mL, graduated in 0.01 mL; 5, 10, and 20-mL capacity.

5.6 *Flasks*, volumetric, 25 and 100-mL capacity.

5.7 *Vibrator*, electric.

5.8 *Vacuum Source*.

5.9 *Evaporator*, vacuum, rotary.

5.10 *Flask*, boiling, round-bottom, short-neck, with 24/40 T joint, 500-mL capacity. Suitable for use with the evaporator (see 5.9).

5.11 *Lamp*, infrared.

5.12 *Burets*, automatic, with integral reservoir, 25-mL capacity.

NOTE 2—Suppliers of stationary phases and supports can be found in Research Report RR:D01-1038, available from ASTM International Headquarters.

6. Reagents and Materials

6.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.³ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

6.2 *Acetone*.

³ *ACS Reagent Chemicals, Specifications and Procedures for Reagents and Standard-Grade Reference Materials*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopoeia and National Formulary*, U.S. Pharmacopoeial Convention, Inc. (USPC), Rockville, MD.

6.3 *Chloroform*.

6.4 *Diatomaceous Earth*⁴—Acid-washed, 60 to 80 mesh and 80 to 100 mesh.

6.5 *Helium*, 99.99 % pure.

6.6 *Methanol*.

6.7 *Methylene Chloride*.

6.8 *Methyl Ethyl Ketone (MEK)*, 99.9 mol %.

6.9 *Methyl Silicone*.⁴

6.10 *1,2,3-Tris(2-Cyanoethoxy) Propane (TCEP)*.⁴

6.11 *Calibration Standards*.

6.11.1 *Benzene*, 99⁺ mol %.

6.11.2 *Isooctane*, 99⁺ mol %.

6.11.3 *n-Nonane*, 99⁺ mol %.

7. Hazards

7.1 Many hydrocarbon solvents are flammable and hazardous; use special precautions when handling them. Of the reagents used in this procedure, methanol, chloroform, methylene chloride, acetone, methyl ethyl ketone, benzene (see 4.1), and *n*-nonane are hazardous.

7.2 Benzene is volatile and highly flammable. Exercise care to prevent accidental ignition. Benzene is also carcinogenic and toxic; acute or chronic poisoning may result from inhalation of benzene vapor, absorption of benzene through the skin, or drinking benzene.

8. Sampling

8.1 Take samples of solvents to be analyzed by this test method using the procedures described in Practice E300.

9. Preparation of Columns

9.1 *Column Packing Preparation*—Prepare the two packing materials, one containing 10 % methyl silicone and the other 25 % TCEP, as follows:

9.1.1 Weigh 45 g of the acid-washed calcined diatomite support 60 to 80 mesh, into a 500-mL flask (see 5.10). Dissolve 5 g of the methyl silicone in approximately 50 mL of chloroform. (**Warning**—Chloroform is a toxic material and inhalation must be avoided.) Pour the methyl silicone–chloroform solution into the flask containing the support. Attach the flask to the evaporator (see 5.9), connect the vacuum, and start the motor. Turn on the infrared lamp and allow the packing to mix thoroughly until dry.

9.1.2 Weigh 75 g of acid-washed pink diatomaceous earth, 80 to 100 mesh, into a 500-mL flask (see 5.10). Dissolve 25 g of TCEP in 200 mL of methanol and pour into the flask containing the support. Attach the flask to the evaporator (see 5.9), connect the vacuum, and start the motor. Turn on the infrared lamp and allow the packing to mix thoroughly until dry, but do not heat the packing above 180 °C.

9.2 *Column Preparation*:

⁴ See Note 2.