



Designation: D5348 – 95 (Reapproved 2019)

Standard Test Method for Determination of the Moisture Content of Sulfonated and Sulfated Oils by Distillation with Xylene¹

This standard is issued under the fixed designation D5348; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the determination of water existing in a sample of sulfonated or sulfated oil, or both, by distilling the sample with a volatile solvent. The test method is applicable only to sulfonated and sulfated oils that do not contain the following: mineral acids, free sulfonic acids, or free sulfuric acid esters; or alkali hydroxides, carbonates or acetates; or alcohol, glycerin, diethylene glycol, acetone, or other water-miscible volatile compounds. This test method was derived from Test Methods **D500**, Sections 4 through 9.

1.2 The values stated in SI units are to be regarded as the standard. The values given in parentheses are provided for information only.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.4 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:²

¹ This test method is under the jurisdiction of ASTM Committee **D31** on Leather and is the direct responsibility of Subcommittee **D31.08** on Fats and Oils. This test method was developed in cooperation with the American Leather Chemists Assn. (Method H 40-1957).

Current edition approved April 1, 2019. Published April 2019. Originally approved in 1993. Last previous edition approved in 2012 as D5348 – 95(2012). DOI: 10.1520/D5348-95R19.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

D500 Test Methods of Chemical Analysis of Sulfonated and Sulfated Oils

3. Significance and Use

3.1 This test method is intended to determine the moisture content of fats, oils, and fatliquors used in the softening and stuffing of leather. The moisture content is measured for the purpose of quality assurance.

4. Apparatus

4.1 The apparatus required consists of a glass flask heated by suitable means and provided with a reflux condenser discharging into a trap and connected to the flask. The connections between the trap and the condenser and flask shall be interchangeable ground joints. The trap serves to collect and measure the condensed water and to return the solvent to the flask. A suitable assembly of the apparatus is illustrated in **Fig. 1**.

4.1.1 *Flask*, 500-mL, of either the short-neck, round-bottom type or the Erlenmeyer type.

4.1.2 *Heat Source*—The source of heat may be either an oil bath (stearic acid, paraffin wax, etc.), or an electric heater provided with a sliding rheostat or other means of heat control.

4.1.3 *Condenser*, a water-cooled glass reflux condenser (**Fig. 1**), having a jacket approximately 400 mm (15³/₄ in.) in length with an inner tube 9.5 to 12.7 mm (³/₈ to ¹/₂ in.) in outside diameter. The end of the condenser to be inserted in the trap shall be ground off at an angle of 30° from the vertical axis of the condenser. When inserted into the trap, the tip of the condenser shall be about 7 mm (¹/₄ in.) above the surface of the liquid in the trap after the distillation conditions have been established. **Fig. 1** shows a conventional sealed-in type of condenser, but any other condenser fulfilling the detailed requirements above may be used.

4.1.4 *Trap*, a trap made of well-annealed glass constructed in accordance with **Fig. 1** and graduated as shown to contain 5 mL at 20°C. It shall be subdivided into 0.1-mL divisions, with each 1-mL line numbered (5 mL at top). The error in any indicated capacity may not be greater than 0.05 mL.