



Designation: D6379 – 21<sup>ε1</sup>



Designation: 436/20

## Standard Test Method for Determination of Aromatic Hydrocarbon Types in Aviation Fuels and Petroleum Distillates—High Performance Liquid Chromatography Method with Refractive Index Detection<sup>1,2</sup>

This standard is issued under the fixed designation D6379; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

<sup>ε1</sup> NOTE—Editorially corrected 9.4.3.1 in October 2021.

### INTRODUCTION

This test method is intended to be technically equivalent to IP 436-20 with an identical title. The ASTM format for test methods has been used, and where possible, equivalent ASTM test methods have replaced the IP or ISO standards.

The test method is intended to be used as one of several possible alternative instrumental test methods that are aimed at quantitative determination of hydrocarbon types in fuels. This does not imply that a correlation necessarily exists between this and any other test method intended to give this information, and it is the responsibility of the user to determine such correlation if necessary.

### 1. Scope\*

1.1 This test method covers a high performance liquid chromatographic test method for the determination of mono-aromatic and di-aromatic hydrocarbon contents in aviation kerosenes and petroleum distillates boiling in the range from 50 °C to 300 °C, such as Jet A or Jet A-1 fuels. The total aromatic content is calculated from the sum of the individual aromatic hydrocarbon-types.

NOTE 1—Samples with a final boiling point greater than 300 °C that contain tri-aromatic and higher polycyclic aromatic compounds are not determined by this test method and should be analyzed by Test Method D6591 or other suitable equivalent test methods.

1.2 This test method is applicable to distillates containing from 0.8 % to 44.0 % by mass mono-aromatic hydrocarbons, 0.23 % to 6.20 % by mass di-aromatic hydrocarbons, and

0.7 % to 50 % by mass total aromatics. Although this method generates results in m/m, results may also be quoted in v/v.

1.3 The precision of this test method has been established for kerosene boiling range distillates containing from 0.40 % to 44.0 % by mass mono-aromatic hydrocarbons, 0.02 % to 6.20 % by mass di-aromatic hydrocarbons, and 0.40 % to 50.0 % by mass total aromatics. If results are quoted in volume, the precision is 0.3 % to 41.4 % by volume mono-aromatics, 0.01 % to 5.00 % by volume di-aromatics, and 0.30 % to 46.3 % by volume total aromatics. As calculated by IP 367-1.

1.4 Compounds containing sulfur, nitrogen, and oxygen are possible interferents. Mono-alkenes do not interfere, but conjugated di- and poly-alkenes, if present, are possible interferents.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.6 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee D02 on Petroleum Products, Liquid Fuels, and Lubricants and is the direct responsibility of Subcommittee D02.04.0C on Liquid Chromatography. The technically equivalent standard as referenced is under the jurisdiction of the Energy Institute Subcommittee SC-G-2.

Current edition approved July 15, 2021. Published September 2021. Originally approved in 1999. Last previous edition approved in 2019 as D6379 – 11 (2019). DOI: 10.1520/D6379-21E01.

<sup>2</sup> This test method has been developed through the cooperative effort between ASTM and the Energy Institute, London. ASTM and IP standards were approved by ASTM and EI technical committees as being technically equivalent but that does not imply both standards are identical.

\*A Summary of Changes section appears at the end of this standard

## 2. Referenced Documents

### 2.1 *ASTM Standards*:<sup>3</sup>

**D4052** Test Method for Density, Relative Density, and API Gravity of Liquids by Digital Density Meter

**D4057** Practice for Manual Sampling of Petroleum and Petroleum Products

**D4177** Practice for Automatic Sampling of Petroleum and Petroleum Products

**D6591** Test Method for Determination of Aromatic Hydrocarbon Types in Middle Distillates—High Performance Liquid Chromatography Method with Refractive Index Detection

### 2.2 *Energy Institute Standards*:<sup>4</sup>

**IP 367–1 (EN ISO 4259 Part 1)** Petroleum and related products – Precision of measurement methods and results – Part 1: Determination of precision data in relation to methods of test

**IP 436** Test method for determination of aromatic hydrocarbon types in aviation fuels and petroleum distillates—High performance liquid chromatography method with refractive index

## 3. Terminology

### 3.1 *Definitions of Terms Specific to This Standard*:

3.1.1 *di-aromatic hydrocarbons (DAHs)*, *n*—compounds that have a longer retention time on the specified polar column than the MAHs.

3.1.2 *mono-aromatic hydrocarbons (MAHs)*, *n*—compounds that have a longer retention time on the specified polar column than the non-aromatic hydrocarbons but a shorter retention time than the di-aromatic hydrocarbons.

3.1.3 *non-aromatic hydrocarbons*, *n*—compounds that have a shorter retention time on the specified polar column than the mono-aromatic hydrocarbons.

3.1.4 *total aromatic hydrocarbons*, *n*—sum of the MAHs and DAHs.

NOTE 2—The elution characteristics of aromatic and non-aromatic compounds on the specified polar column have not been specifically determined for this test method. Published and unpublished data indicate the major constituents for each hydrocarbon type as follows: (1) Non-aromatic hydrocarbons: acyclic and cyclic alkanes (paraffins and naphthenes), mono-alkenes (if present). (2) MAHs: benzenes, tetralins, indanes, thiophenes, conjugated poly-alkenes. (3) DAHs: naphthalenes, biphenyls, indenes, fluorenes, acenaphthenes, benzothiophenes.

## 4. Summary of Test Method

4.1 The test portion is diluted 1:1 with the mobile phase, such as heptane, and a fixed volume of this solution injected into a high performance liquid chromatograph fitted with a set of polar columns. These columns have little affinity for the non-aromatic hydrocarbons and exhibits a pronounced selectivity for aromatic hydrocarbons. As a result of this selectivity,

the aromatic hydrocarbons are separated from the non-aromatic hydrocarbons into distinct bands in accordance with their ring structure, that is, MAHs and DAHs.

4.2 The column is connected to a refractive index detector that detects the components as they elute from the column. The electronic signal from the detector is continually monitored by a data processor. The amplitudes of the signals (peak areas) from the sample aromatics are compared with those obtained from previously-run calibration standards in order to calculate the percent m/m MAHs and DAHs in the sample. The sum of the MAHs and DAHs is reported as the total aromatic content (percent m/m) of the sample. Although this method generates results in m/m, results can also be quoted in percent v/v either by calibrating in v/v or by converting m/m to v/v by using the densities of the sample and standards.

## 5. Significance and Use

5.1 Accurate quantitative information on aromatic hydrocarbon types can be useful in determining the effects of petroleum processes on production of various finished fuels. This information can also be useful for indicating the quality of fuels and for assessing the relative combustion properties of finished fuels.

## 6. Apparatus

6.1 *High Performance Liquid Chromatograph (HPLC)*—Any high performance liquid chromatograph capable of pumping the mobile phase at flow rates between 0.5 mL/min and 1.5 mL/min with a precision better than 0.5 % and a pulsation of <1 % full scale deflection under the test conditions described in Section 9. See Fig. 1.

6.2 *Sample Injection System*—The sample injection system capable of injecting 5  $\mu$ L (nominal) of sample solution with a repeatability better than 2 %.

6.2.1 An equal and constant volume of the calibration and sample solutions shall be injected into the chromatograph. Both manual and automatic sample injection systems (using either complete or partial filling of the sample loop) will, when used correctly, meet the repeatability requirements laid down in 6.2. When using the partial loop filling mode, it is recommended that the injection volume should be less than half the total loop volume. For complete filling of the loop, best results are obtained by overfilling the loop at least six times.

6.2.2 Sample injection volumes other than 5  $\mu$ L (typically in the range from 3  $\mu$ L to 20  $\mu$ L) may be used provided they meet the requirements laid down for injection repeatability (see 6.2), refractive index sensitivity and linearity (see 9.4 and 10.1), and column resolution (see 9.4)

6.3 *Sample Filter (Optional)*—A microfilter of porosity 0.45  $\mu$ m or less, which is chemically-inert towards hydrocarbon solvents, is recommended for the removal of particulate matter from the sample solutions.

6.4 *Column System*—Any stainless steel HPLC column(s) packed with an approved amino-bonded (or polar amino/cyano-bonded) silica stationary phase is suitable. The column(s) used shall satisfy the resolution requirements laid down in 9.4.3. Column lengths from 150 mm to 300 mm with an

<sup>3</sup> For referenced ASTM standards, visit the ASTM website, [www.astm.org](http://www.astm.org), or contact ASTM Customer Service at [service@astm.org](mailto:service@astm.org). For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

<sup>4</sup> Available from Energy Institute, 61 New Cavendish St., London, W1G 7AR, U.K., <http://www.energyinst.org>.