



Designation: D7164 – 21

Standard Practice for On-line/At-line Heating Value Determination of Gaseous Fuels by Gas Chromatography¹

This standard is issued under the fixed designation D7164; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This practice is for the determination of heating value in high methane content gaseous fuels such as natural gas using an on-line/at-line gas chromatograph.

1.2 *Units*—The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.4 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:²

- D1070 Test Methods for Relative Density of Gaseous Fuels
- D1945 Test Method for Analysis of Natural Gas by Gas Chromatography
- D1946 Practice for Analysis of Reformulated Gas by Gas Chromatography
- D3588 Practice for Calculating Heat Value, Compressibility Factor, and Relative Density of Gaseous Fuels
- D3764 Practice for Validation of the Performance of Process Stream Analyzer Systems
- D4150 Terminology Relating to Gaseous Fuels
- D4626 Practice for Calculation of Gas Chromatographic Response Factors

¹ This practice is under the jurisdiction of ASTM Committee D03 on Gaseous Fuels and is the direct responsibility of Subcommittee D03.12 on On-Line/At-Line Analysis of Gaseous Fuels.

Current edition approved April 1, 2021. Published May 2021. Originally approved in 2005. Last previous edition approved in 2015 as D7164 – 10(2015). DOI: 10.1520/D7164-21.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

- D5287 Practice for Automatic Sampling of Gaseous Fuels
 - D5503 Practice for Natural Gas Sample-Handling and Conditioning Systems for Pipeline Instrumentation (Withdrawn 2017)³
 - D6122 Practice for Validation of the Performance of Multi-variate Online, At-Line, Field and Laboratory Infrared Spectrophotometer, and Raman Spectrometer Based Analyzer Systems
 - D6299 Practice for Applying Statistical Quality Assurance and Control Charting Techniques to Evaluate Analytical Measurement System Performance
 - D6621 Practice for Performance Testing of Process Analyzers for Aromatic Hydrocarbon Materials
 - D7833 Test Method for Determination of Hydrocarbons and Non-Hydrocarbon Gases in Gaseous Mixtures by Gas Chromatography
 - E260 Practice for Packed Column Gas Chromatography
 - E594 Practice for Testing Flame Ionization Detectors Used in Gas or Supercritical Fluid Chromatography
 - E1510 Practice for Installing Fused Silica Open Tubular Capillary Columns in Gas Chromatographs
- ### 2.2 ISO Standards⁴
- ISO 6976 Natural Gas — Calculation of Calorific Values, Density, Relative Density and Wobbe Indices From Composition
 - ISO 7504 Gas Analysis-Vocabulary
- ### 2.3 Gas Processors Association Standards:⁵
- GPA 2172 Calculation of Gross Heating Value, Relative Density, Compressibility and Theoretical Hydrocarbon Liquid Content for Natural Gas Mixtures for Custody Transfer

3. Terminology

3.1 For definitions of general terms used in D03 Gaseous Fuels standards, refer to Terminology D4150.

³ The last approved version of this historical standard is referenced on www.astm.org.

⁴ Available from International Organization for Standardization (ISO), 1, ch. de la Voie-Creuse, Case postale 56, CH-1211, Geneva 20, Switzerland, <http://www.iso.ch>.

⁵ Available from Gas Processors Association (GPA), 66 American Plaza, Suite 700, Tulsa, OK 74135, <http://www.gpaglobal.org>.

4. Summary of Practice

4.1 A representative sample of the gaseous fuel is extracted from a process pipe or a pipeline and is transferred in a timely manner to an analyzer sampling system. After appropriate conditioning steps that maintain the sample's integrity are completed, a precise volume of sample is injected onto an appropriate gas chromatographic column. Excess extracted process or pipeline sample is vented to atmosphere, a flare header, or is returned to the process in accordance with applicable economic and environmental requirements and regulations.

4.2 Sample constituents are separated in the column to elute individually for identification and quantification by the detector and its data handling system. Heating value is calculated using the results of the compositional analysis and the calculations contained in Practice [D3588](#) or other standard algorithms.

4.3 Calibration, maintenance, and performance protocols provide a means to validate the analyzer operation.

5. Significance and Use

5.1 On-line, at-line, in-line, and other near-real time monitoring systems that measure fuel gas characteristics such as the heating value are prevalent in the natural gas and fuel gas industries. The installation and operation of particular systems vary on the specific objectives, process type, regulatory requirements, and internal performance requirements needed by the user. This protocol is intended to provide guidelines for standardized start-up procedures, operating procedures, and quality assurance practices for on-line, at-line, in-line, and other near-real time heating value monitoring systems.

6. Apparatus

6.1 *Instrument*—Any instrument of standard manufacture, with hardware necessary for interfacing to a natural gas or other fuel gas pipeline and containing all the features necessary for the intended application(s), can be used.

6.1.1 *Chromatographic-based Systems*—The chromatographic parameters employed generally should be capable of obtaining a relative retention time repeatability of 0.05 min (3 s) for duplicate measurements. Instrumentation should satisfy or exceed other chromatographic and analytic performance characteristics for accuracy and precision for the intended application without encountering unacceptable interference or bias. In addition, components in contact with sample streams such as tubing and valving must be constructed of suitable inert materials to ensure constituents in the fuel stream do not degrade these components or alter the composition of the sampled gas. Additional information related to analyzing gaseous fuels using gas chromatography can be found in Test Method [D1945](#), Practice [D1946](#), and Test Method [D7833](#).

6.2 *Sample Probes/Sample Extraction*—The location and orientation of sampling components are critical for ensuring that a representative sample is analyzed. The locations and orientation of sampling components should be selected based upon sound analytic and engineering considerations. Sampling practices for gaseous fuels can be found in Practice [D5287](#).

6.3 *Sample Inlet System*—The siting and installation of an at-line or on-line monitor is critical for collecting representative information on heating value content. Factors that should be considered in siting an instrument include ease of calibration, ease of access for repair or maintenance, sample uniformity at the sampling point, appropriateness of samples from a sampling location, ambient conditions, and of course safety issues. An automated gas sampling valve is required in many applications. All sampling system components in contact with the fuel stream must be constructed of inert or passivated materials. Care must be taken to ensure that the extracted sample is maintained in the gaseous phase throughout pressure reduction and transport throughout the expected ranges of ambient and pipeline temperatures and pressure. The addition of heat at the point of pressure reduction and along the sample line to the analyzer may be required to ensure that the entire sample is maintained in the gas phase during all expected operating conditions and gas compositions. When selecting sample conditioning equipment, equations of state should be used to determine whether pressure must be reduced in multiple stages to prevent a change of phase. The need for heat tracing and the extent to which it is required will be site specific. In general, considerations impacting heat tracing decisions include sample compositions and the expected variations, ambient temperature fluctuations, operating pressures, and anticipated pressure differentials in sample system components. Sample filtration should be utilized as required to remove particulate matter from the extracted sample and to prevent any contaminating liquids such as glycol, compressor oils, and liquid water, which may form in the event of heat-tracing failure, from reaching the chromatograph. The sampling frequency relative to the process bandwidth is critical to ensuring that the reported analytical results adequately represent the process being monitored. The Nyquist-Shannon sampling criterion of a sampling frequency that exceeds twice the process bandwidth can be used to establish a minimum analytical cycle time. Sample handling and conditioning system practices can be found in Practice [D5503](#).

6.3.1 *Carrier and Detector Gas Control*—Constant flow carrier and detector gas control is critical for optimum and consistent analytical performance. Control is achieved by use of pressure regulators, mass flow controllers, and fixed flow restrictors. Temperature control is generally vital for ensuring consistent operation of these devices. The gas flow is measured by appropriate means and adjusted as necessary. Most chromatographs automatically control their gas flow rates which reduces a user's responsibility for ensuring the proper gas pressure is provided to the chromatograph. Should a chromatograph require external gas flow control, mass flow controllers, capable of maintaining gas flow constant to $\pm 1\%$ at the flow rates necessary for optimal instrument performance, are generally used.

6.3.2 *Detectors*—A thermal conductivity detector (TCD) is commonly used. Other detectors, such as the flame ionization detector (FID), Practice [E594](#), can be used but should at least meet TCD linearity, sensitivity, and selectivity in the selected application.