



Standard Test Method for Bromide Ion in Water¹

This standard is issued under the fixed designation D1246; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the U.S. Department of Defense.

1. Scope*

1.1 This test method is applicable to the measurement of bromide ion in water, ground water, and drinking water.

1.2 Samples containing 0.5 mg/L to 1000 mg/L of bromide may be analyzed by this test method. The concentration range may be extended by the dilution of an appropriate aliquot.

1.3 The precision and bias statements were determined on natural and ground waters. It is the responsibility of the analyst to assure the validity of this test method for untested matrices.

1.4 A titrimetric and two colorimetric test methods for iodide and bromide were discontinued. Refer to [Appendix X1](#) for historical information.

1.5 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.6 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:²

[D1066 Practice for Sampling Steam](#)

[D1129 Terminology Relating to Water](#)

[D1193 Specification for Reagent Water](#)

[D2777 Practice for Determination of Precision and Bias of Applicable Test Methods of Committee D19 on Water](#)

[D3370 Practices for Sampling Water from Closed Conduits](#)

[D4127 Terminology Used with Ion-Selective Electrodes](#)

[D5810 Guide for Spiking into Aqueous Samples](#)

¹ This test method is under the jurisdiction of ASTM Committee D19 on Water and is the direct responsibility of Subcommittee D19.05 on Inorganic Constituents in Water.

Current edition approved June 15, 2016. Published June 2016. Originally approved in 1952. Last previous edition approved in 2010 as D1246 – 10. DOI: 10.1520/D1246-16.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

[D5847 Practice for Writing Quality Control Specifications for Standard Test Methods for Water Analysis](#)

3. Terminology

3.1 Definitions:

3.1.1 For definitions of terms used in this standard, refer to Terminologies [D1129](#) and [D4127](#).

4. Summary of Test Method

4.1 Bromide ion is measured potentiometrically using a bromide ion-selective electrode in conjunction with a single-junction, sleeve-type reference electrode. Potentials are read on a pH meter having an expanded millivolt scale capable of being read to the nearest 0.1 mV, or a selective ion meter having a direct concentration scale for bromide. For less precise work, a pH meter having a millivolt scale capable of being read to the nearest 1.0 mV is adequate, however, no supporting data are available.

4.2 The electrodes are calibrated in known bromide solutions and the concentrations of unknowns are determined in solutions with the same background. In most cases, addition of an ionic strength adjustor to both standards and samples is sufficient to maintain constant background ionic strength. For samples above 0.1 M in ionic strength, prepare standard solutions similar to the sample composition.

5. Significance and Use

5.1 By analysis for bromide in water, wastewater, and brackish waters, it is possible to evaluate the origin of the water, its potential as a source of bromide, and its condition with regard to pollution.

6. Interferences

6.1 Strongly reducing solutions and solutions containing ions which form insoluble silver salts may coat the electrode membrane. These may be removed by polishing the membrane surface. Sulfide ion and cyanide ion both poison the electrode, and should be removed (see Section 11).

6.2 Halide ions form complexes with some metals. Since the electrode responds only to free bromide ions, the presence of any complexing agents lowers the measured concentrations. Concentrations of free metal ions causing a 10 % error of a

*A Summary of Changes section appears at the end of this standard

bromide concentration of 8.1 mg/L are bismuth⁺³, 80 mg/L; cadmium⁺², 100 mg/L; lead⁺², 1600 mg/L; tin⁺², 2400 mg/L; and thallium⁺³, 4 mg/L.

6.3 Chloride ion and hydroxide ion do not interfere when present in the concentrations of up to 400 and 30 000 times the bromide concentration, respectively. There will be no interference from ammonia when present in concentrations twice that of bromide, nor from thiosulfate 20 times as concentrated as bromide. Iodide is an interference at a concentration ratio as low as 2×10^{-4} . Mercury should be absent from samples.

7. Apparatus

7.1 *pH Meter*, capable of reading to 0.1 mV, or a selective-ion meter.

7.2 *Bromide Ion-Selective Electrode*.

7.3 *Sleeve-Type Single-Junction Reference Electrode*, filled with manufacturer's filling solution.

7.4 *Mixer*, magnetic, with a TFE fluorocarbon-coated stirring bar.

7.5 *Laboratory Glassware*.

8. Reagents

8.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.³ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

8.2 *Purity of Water*—Unless otherwise indicated, reference to water shall be understood to mean reagent water conforming to Specification **D1193**, Type I. Other reagent water types may be used, provided it is first ascertained that the water is of sufficiently high purity to permit its use without adversely affecting the bias and precision of the test method. Type II water was specified at the time of round robin testing of this test method.

8.3 *Ionic Strength Adjustor* (42.5 g/100 mL NaNO₃)—In a 100-mL volumetric flask, dissolve 42.5 g of sodium nitrate (NaNO₃) in water and dilute to volume.

8.4 *Nickel Sulfate Solution, 1 M*—In a 100 mL volumetric flask dissolve 26.3 g of nickel sulfate hexahydrate (NiSO₄ · 6H₂O) in water and dilute to volume.

8.5 *Sodium Bromide Standard Solution, 1000 mg/L*—In a 1 L volumetric flask dissolve 1.288 g dried sodium bromide in water and dilute to volume. Alternatively, certified bromide stock solutions of appropriate known purity are commercially available through chemical supply vendors and may be used.

³ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmaceutical Convention, Inc. (USPC), Rockville, MD.

8.6 *Sodium Bromide Standard Solutions*, (100, 10, and 1 mg/L)—Dilute 1 volume of the 1000 mg/L bromide standard with 9 volumes of water to prepare the 100 mg/L standard. By further 1+9 serial dilutions, prepare the 10 and 1 mg/L standards.

9. Sampling

9.1 Collect the samples in accordance with Practice **D1066** and Practices **D3370**, as applicable.

10. Calibration and Standardization

10.1 To 100 mL of the 1, 10, 100, and 1000 mg/L NaBr standards add 2 mL of the ionic strength adjustor (ISA). If any samples require treatment for interferences, prepare standards with the same background.

10.2 Connect and fill electrodes in accordance with manufacturer's instructions.

10.3 Transfer the 1 mg/L standard-ISA mixture to a 150 mL beaker and stir gently using the magnetic mixer. Immerse the electrodes in the solution and wait 2 min for the potential to stabilize. Record the value.

10.4 Rinse electrodes thoroughly and repeat for the 10, 100, and 1000 mg/L standard-ISA mixtures. Wait 2 min and record the potential.

10.5 Follow manufacturer instructions for calibrating selective-ion meters with a direct reading of concentration capabilities. For pH meters, generate a calibration curve by creating a calibration curve on semilogarithmic graph paper or by creating a plot of mV potential versus log(10) concentration. Note that volume corrections are incorporated into the calibration, so that samples analyzed according to Section **11** of this test method can be read directly.

11. Procedure

11.1 To any sample containing sulfide or cyanide ion, add 0.1 mL nickel sulfate solution to 100 mL sample.

NOTE 1—This concentration of nickel sulfate will react with 58 mg/L sulfide and 117 mg/L cyanide.

11.2 Transfer 100 mL sample to a 150 mL beaker and add 2 mL ionic strength adjustor. Stir thoroughly for 1 min using the magnetic mixer.

11.3 Immerse the electrodes in the sample-ISA mixture and wait 2 min for the potential to stabilize. Record the value.

11.4 Read bromide concentration of the sample, in mg/L, directly from the meter. Note that volume corrections are incorporated into the calibration.

12. Precision and Bias⁴

12.1 *Precision*—The overall and single-operator precision of this test method may be expressed as follows:

⁴ Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR:D19-1078. Contact ASTM Customer Service at service@astm.org.