

Standard Guide to Charge Control and Charge Referencing Techniques in X-Ray Photoelectron Spectroscopy¹

This standard is issued under the fixed designation E1523; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This guide acquaints the X-ray photoelectron spectroscopy (XPS) user with the various charge control and charge shift referencing techniques that are and have been used in the acquisition and interpretation of XPS data from surfaces of insulating specimens and provides information needed for reporting the methods used to customers or in the literature.

1.2 This guide is intended to apply to charge control and charge referencing techniques in XPS and is not necessarily applicable to electron-excited systems.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

2. Referenced Documents

2.1 ASTM Standards:²

- E673 Terminology Relating to Surface Analysis (Withdrawn 2012)³
- E902 Practice for Checking the Operating Characteristics of X-Ray Photoelectron Spectrometers (Withdrawn 2011)³
- E1078 Guide for Specimen Preparation and Mounting in Surface Analysis
- E1829 Guide for Handling Specimens Prior to Surface Analysis

3. Terminology

3.1 *Definitions*—See Terminology E673 for definitions of terms used in XPS.

3.2 Symbols:		
BE	Binding energy, in eV	
BE _{corr}	Corrected binding energy, in eV	
BE _{meas}	Measured binding energy, in eV	
BE _{ref}	Reference binding energy, in eV	
BE _{meas, ref}	Measured Binding energy, in eV, of a reference line	
FWHM	Full width at half maximum amplitude of a peak in the	
	photoelectron spectrum above the background, in eV	
XPS	X-ray photoelectron spectroscopy	
Δ_{corr}	Correction energy, to be added to measured binding	
	energies for charge correction, in eV	

4. Overview of Charging Effects

4.1 For insulating specimen surfaces, the emission of photoelectrons following X-ray excitation may result in a temporary (or sometimes persistent) buildup of a positive surface charge caused by the photoelectric effect. Its insulating nature prevents the compensation of the charge buildup by means of electron conduction from the sample holder. This positive surface charge changes the surface potential thereby shifting the measured energies of the photoelectron peaks to higher binding energy. This binding energy shift may reach a nearly steady-state value of between 2 and 5 eV for spectrometers equipped with nonmonochromatic X-ray sources. The surface potential charge and the resulting binding energy shift is, generally, larger for spectrometers equipped with monochromatic X-ray sources because of the, generally, lower flux of low-energy electrons impinging on the specimen surface. This lower flux arises because focused, monochromatic X-ray beams irradiate only a portion of the specimen and not other nearby surfaces (for example, the specimen holder) that are sources of low-energy electrons. The absence of an X-ray window in many monochromatic X-ray sources (or a greater distance of the specimen from the X-ray window) also eliminates another source of low-energy electrons.

4.2 The amount of induced surface charge, its distribution across the specimen surface, and its dependence on experimental conditions are determined by several factors including specimen composition, homogeneity, magnitude of surface conductivity, total photoionization cross-section, surface topography, spatial distribution of the exciting X-rays, and

¹ This guide is under the jurisdiction of ASTM Committee E42 on Surface Analysis and is the direct responsibility of Subcommittee E42.03 on Auger Electron Spectroscopy and X-Ray Photoelectron Spectroscopy.

Current edition approved June 1, 2015. Published June 2015. Originally approved in 1993. Last previous edition approved in 2009 as E1523 – 09. DOI: 10.1520/E1523-15.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

 $^{^{3}\,\}text{The}$ last approved version of this historical standard is referenced on www.astm.org.

availability of neutralizing electrons. Charge buildup is a well-studied $(1, 2)^4$, three dimensional phenomenon that occurs along the sample surface and into the material. The presence of particles on or different phases in the specimen surface may result in an uneven distribution of charge across the surface, a phenomenon known as differential charging. Charge buildup may also occur at phase boundaries or interface regions within the depth of the sample that is impacted by X-ray radiation.

4.3 Several techniques have been developed for the purpose of controlling charge buildup and the subsequent changes in surface potential in order to obtain meaningful and reproducible data from insulating specimens. These techniques are employed during the data acquisition and are discussed in 7.2.

4.4 Several techniques have been developed for the purpose of correcting the binding energy shifts that result from surface charging. These corrections are performed after the data has been accumulated and are discussed in 7.3.

4.5 The use of the various charge control or charge referencing techniques described in this guide may depend on the available instrument as well as the specimen being analyzed.

4.6 Specimens with non-insulating surfaces are those with a high enough electron conductivity to dynamically compensate the electron loss caused by the photoelectric effect; they neither require control of the surface charge buildup nor charge reference corrections. It is important to distinguish the shifts due to the temporary charge build caused by the photoelectric effect from intrinsic charging effects. Intrinsic effects, such as the accumulation of charge at an interface during film growth, influence the nature of spectra obtained and the BEs measured, but are part of the sample (3). It is also possible that the impinging of the X-ray changes the charge distribution by means of volatilization of certain chemical species or the creation or charge centers. Such specimens may never achieve steady-state potentials. Although artifact to the process of measurement, those changes become part of the sample and are not necessarily to be corrected or compensated by the methods described in 7.2 and 7.3.

4.7 Major advances in the ability to control sample charging and to stabilize surface potential were made in the late 1990s including the ability to achieve charge control for small area analysis (4). These approaches usually involve the use of electron flood guns and some additional methods (ions or magnetic fields) to control localized surface charge (5, 6). As a result of these advances it is now possible to collect high quality reproducible data on many systems. However, these advances do not remove all of the challenges for optimizing the conditions for analysis for complex samples or interpreting the data.

4.8 Although changes in surface potential during XPS analysis and other charging effects are usually viewed as problems to be avoided, such phenomena can be used to extract important information about specimens (**7-9**).

5. Significance and Use

5.1 The acquisition of chemical information from variations in the energy position of peaks in the XPS spectrum is of primary interest in the use of XPS as a surface analytical tool. Surface charging acts to shift spectral peaks independent of their chemical relationship to other elements on the same surface. The desire to eliminate the influence of surface charging on the peak positions and peak shapes has resulted in the development of several empirical methods designed to assist in the interpretation of the XPS peak positions, determine surface chemistry, and allow comparison of spectra of conducting and non-conducting systems of the same element. It is assumed that the spectrometer is generally working properly for non-insulating specimens (see Practice E902).

5.2 Although highly reliable methods have now been developed to stabilize surface potentials during XPS analysis of most materials (5, 6), no single method has been developed to deal with surface charging in all circumstances (10, 11). For insulators, an appropriate choice of any control or referencing system will depend on the nature of the specimen, the instruments, and the information needed. The appropriate use of charge control and referencing techniques will result in more consistent, reproducible data. Researchers are strongly urged to report both the control and referencing techniques that have been used, the specific peaks and binding energies used as standards (if any), and the criteria applied in determining optimum results so that the appropriate comparisons may be made.

6. Apparatus

6.1 One or more of the charge compensation techniques mentioned in this guide may be employed in virtually any XPS spectrometer.

6.2 Some of the techniques outlined require special accessory apparatus, such as electron flood sources or a source for evaporative deposition.

6.3 Certain specimen mounting procedures, such as mounting the specimen under a fine metal mesh (12), can enhance electrical contact of the specimen with the specimen holder, or reduce the amount of surface charge buildup. This and other methods of specimen mounting to reduce static charge are described in detail in Guide E1078 and Guide E1829.

7. Procedures

7.1 The methods described here involve charge control (the effort to control the buildup of charge at a surface or to minimize its effect), charge referencing (the effort to determine a reliable binding energy despite buildup of charge), or some combination of the two. For charge control, peak shape is the most important parameter to consider. A constant and relatively uniformly surface potential provides the conditions needed to obtain reproducible data and optimum peak shape. Correcting the peak position is accomplished separately using an appropriate charge referencing technique. In some circumstances, the Auger parameter can provide chemical information without the need to resort to surface potential corrections.

⁴ The boldface numbers in parentheses refer to a list of references at the end of this standard.

7.2 A variety of different methods is used to either enhance conductivity to minimize charge buildup during XPS analysis or to control the surface potential by other methods. These methods employed to control the surface potential in insulating specimens are listed in Table 1 in approximate order of frequency of use (more frequently used first) and summarized below:

7.2.1 Methods for Controlling the Sample Surface Potential:

7.2.1.1 *Electron Flood Gun* (13-16)—Use of low-energy electron flood guns to stabilize the surface potential of insulators examined by XPS (14), in particular when monochromatized X-rays are employed. Optimum operating conditions, for example, filament position, electron energy, and electron current, depend upon the orientation of the electron flood gun with respect to the specimen and upon the particular design of the electron flood gun and must, in general, be determined by the user. Use low-electron energies (usually 10 eV or less) to maximize the neutralization effect and reduce the number of electron bombardment-induced reactions. A metal screen placed on or above the specimen can help (17, 18).

7.2.1.2 Low Energy Ion Source—Recent work indicates that portions of an insulator surface can be negatively charged, even when some areas exposed to X-rays are charged positively (19). Such effects appear to be particularly important for focused X-ray beam systems, where the X-rays strike only a relatively small portion of the specimen. In these circumstances, the use of a low-energy positive-ion source, in addition to an electron source, helps stabilize (and make more uniform) the surface potential of the specimen. Several commercial XPS now effectively combine electrons and ions to achieve uniform surface potentials for many types of insulators.

7.2.1.3 Ultraviolet Flood Lamp (20)—Ultraviolet radiation can also produce low-energy electrons (for example, from the specimen holder) that may be useful in neutralizing specimen charging and stabilizing the surface potential.

7.2.1.4 *Biasing*—Applying a low-voltage bias (-10 to +10 V) to the specimen and observing the changes in the binding energies of various peaks can be used to learn about the electrical contact of a specimen (or parts of a specimen) with the specimen holder. Peaks in the XPS spectrum that shift when the bias is applied are from conducting regions of the specimen. Other peaks from insulating regions may not shift nearly as much or at all and can be interpreted accordingly. This method can sometimes verify that the peaks being used for charge referencing (for example, gold 4f or carbon 1s) are

TABLE 1 Methods Used to Stabilize or Control Surface Potential During XPS Analysis

Approach/Method	Section
Controlling the Sample Surface Potential	7.2.1
Electron Flood Gun	7.2.1.1
Low Energy Ion Source	7.2.1.2
Ultraviolet Flood Lamp	7.2.1.3
Biasing	7.2.1.4
Isolation from Ground	7.2.1.5
Minimizing Charge Accumulation	7.2.2
Grounding and Enhanced Conduction Path	7.2.2.1
Specimen Heating	7.2.2.2

behaving in the same manner as the peaks of interest from the specimen (12, 20, 21). For non-uniform or composite (nonconducting or partially conducting) specimens, a variety of charge shifts may be observed upon biasing. This may provide useful information about the sample and indicate a need to more carefully connect the specimen to ground or to isolate the sample from ground. Sometimes all data for some specimens are collected with a bias applied (see also 7.4).

7.2.1.5 *Isolation from Ground*—For some materials, or mixtures of materials with different electrical conductivity, differential charging can occur. This phenomenon can be used to obtain information about the sample (4, 22) and can sometimes be minimized (and a more uniform sample potential can be achieved) by isolating the specimen from ground. In some circumstances an electron flood gun is more effective in controlling the surface potential when the sample is isolated from ground.

7.2.2 *Methods for Minimizing Charge Accumulation*— These methods attempt to stabilize the surface potential by minimizing the charge buildup or potential change by lowering sample resistance to ground or the spectrometer mount.

7.2.2.1 Grounding and Enhanced Conduction Path— Surrounding of insulating materials with a conducting material has been a common approach to minimizing the charge build up on samples. This can mean masking a solid sample with a conducting aperture, grid, or foil or mounting particles on a conducting foil or tape (2).

7.2.2.2 *Specimen Heating*—For a limited number of specimens, heating can increase the electrical conductivity of the specimen, thus decreasing charging (2).

7.3 Binding Energy Reference Methods—A variety of methods (as listed in Table 2 and described below) have been used to determine the amount of binding energy shift resulting from surface charging in insulating specimens. Each of these methods is based on the assumption that differential charging (along the surface or within the sample) is not present to a significant degree. If significant differential charging is found to occur or thought to be present, it may be necessary to alter the method of charge (potential) control.

7.3.1 Adventitious Carbon Referencing (12, 13, 20, 23-27)—Unless specimens are prepared for analysis under carefully controlled atmospheres, the surface, generally, is coated by adventitious contaminants. Once introduced into the spectrometer, further specimen contamination can occur by the adsorption of residual gases, especially in instruments with oil diffusion pumps. These contamination layers can be used for referencing purposes if it is assumed that they truly reflect the steady-state static charge exhibited by the specimen surface and that they contain an element with a peak of known binding energy. Carbon is most commonly detected in adventitious

Approach/Method	Section
Adventitious Carbon Referencing	7.3.1
Internal Referencing	7.3.2
Substrate Referencing	7.3.3
Gold Deposition	7.3.4
Implantation with Inert Gases	7.3.5