



Standard Guide for Use of Radiation-Sensitive Indicators¹

This standard is issued under the fixed designation ISO/ASTM 51539; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision.

1. Scope

1.1 This guide covers procedures for using radiation-sensitive indicators (referred to hereafter as *indicators*) in radiation processing. These indicators may be labels, papers, inks or packaging materials which undergo a visual change when exposed to ionizing radiation **(1-5)**.²

1.2 The purpose for using indicators is to determine visually whether or not a product has been irradiated, rather than to measure different dose levels.

1.3 Indicators are not dosimeters and shall not be used as a substitute for proper dosimetry. Information about dosimetry systems for radiation processing is provided in other ASTM and ISO/ASTM documents (see ISO/ASTM Guide 51261).

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced documents

2.1 ASTM Standards:³

E170 Terminology Relating to Radiation Measurements and Dosimetry

2.2 ISO/ASTM Standards:³

51204 Practice for Dosimetry in Gamma Irradiation Facilities for Food Processing

51261 Practice for Calibration of Routine Dosimetry Systems for Radiation Processing

51431 Practice for Dosimetry in Electron Beam and X-ray

(Bremsstrahlung) Irradiation Facilities for Food Processing

51608 Practice for Dosimetry in an X-ray (Bremsstrahlung) Facility for Radiation Processing

51649 Practice for Dosimetry in an Electron Beam Facility for Radiation Processing at Energies between 300 keV and 25 MeV

51702 Practice for Dosimetry in Gamma Facility for Radiation Processing

51939 Practice for Blood Irradiation Dosimetry

51940 Guide for Dosimetry for Sterile Insect Release Programs

2.3 *International Commission on Radiation Units and Measurements (ICRU) Reports:*⁴

ICRU Report 85 Fundamental Quantities and Units for Ionizing Radiation

3. Terminology

3.1 Definitions:

3.1.1 *dosimetry system*—system used for determining absorbed dose, consisting of dosimeters, measurement instruments and their associated reference standards, and procedures for the system's use.

3.1.2 *indicator*—see *radiation-sensitive indicator*.

3.1.3 *process load*—volume of material with a specified product loading configuration irradiated as a single entity.

3.1.4 *radiation-sensitive indicator*—material such as coated or impregnated adhesive-backed substrate, ink, coating or other materials which may be affixed to or printed on the process loads, and which undergoes a visual change when exposed to ionizing radiation.

3.1.4.1 *Discussion*—This visual change may be as observed with the unaided eye or with an eye aid such as an optical filter.

3.1.4.2 *Discussion*—In the past, radiation-sensitive indicators have been referred to by different names **(1-5)** including *label dosimeters*, *label indicators*, *indicator labels*, *dose indicators*, and *radiation monitoring labels*. ASTM Subcommittee E10.01 considers a *label dosimeter* to be fundamentally different from an *indicator*, and intends to define a label dosimeter as a type of routine dosimeter (that is, quantitative)

¹ This guide is under the jurisdiction of ASTM Committee E61 on Radiation Processing and is the direct responsibility of Subcommittee E61.04 on Specialty Application, and is also under the jurisdiction of ISO/TC 85/WG 3.

Current edition approved May 27, 2013. Published July 2013. Originally published as ASTM E 1539–93. Last previous ASTM edition E 1539–98¹. ASTM E 1539–93 was adopted by ISO in 1998 with the intermediate designation ISO 15564:1998(E). The present International Standard ISO/ASTM 51539:2005(2013)(E) replaces ISO 15544 and is a reapproval of the last previous edition ISO/ASTM 51539:2005(E).

² The boldface numbers in parentheses refer to the bibliography at the end of this guide.

³ For referenced ASTM and ISO/ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For Annual Book of ASTM Standards volume information, refer to the standard's Document Summary page on the ASTM website.

⁴ Available from the International Commission on Radiation Units and Measurements, 7910 Woodmont Ave., Suite 800, Bethesda, MD 20814, USA.