



IEC 60601-2-23

Edition 3.0 2011-02

# INTERNATIONAL STANDARD

# NORME INTERNATIONALE

Medical electrical equipment –  
**Part 2-23: Particular requirements for the basic safety and essential performance  
of transcutaneous partial pressure monitoring equipment**

Appareils électromédicaux –  
**Partie 2-23: Exigences particulières pour la sécurité de base et les performances  
essentielles des appareils de surveillance de la pression partielle transcutanée**

INTERNATIONAL  
ELECTROTECHNICAL  
COMMISSION

COMMISSION  
ELECTROTECHNIQUE  
INTERNATIONALE

PRICE CODE  
CODE PRIX

X

ICS 11.040.55

ISBN 978-2-88912-370-4

## CONTENTS

FOREWORD.....	4
INTRODUCTION.....	7
201.1    Scope, object and related standards .....	8
201.2    Normative references.....	10
201.3    Terms and definitions.....	10
201.4    General requirements .....	11
201.5    General requirements for testing of ME EQUIPMENT .....	11
201.6    Classification of ME EQUIPMENT and ME SYSTEMS.....	12
201.7    ME EQUIPMENT identification, marking and documents .....	12
201.8    Protection against electrical HAZARDS from ME EQUIPMENT .....	13
201.9    Protection against MECHANICAL HAZARDS of ME EQUIPMENT and ME SYSTEMS .....	14
201.10   Protection against unwanted and excessive radiation HAZARDS .....	14
201.11   Protection against excessive temperatures and other HAZARDS .....	14
201.12   Accuracy of controls and instruments and protection against hazardous outputs .....	18
201.13   HAZARDOUS SITUATIONS and fault conditions .....	23
201.14   PROGRAMMABLE ELECTRICAL MEDICAL SYSTEMS (PEMS).....	23
201.15   Construction of ME EQUIPMENT .....	23
201.16   ME SYSTEMS .....	23
201.17   Electromagnetic compatibility of ME EQUIPMENT and ME SYSTEMS .....	24
202      Electromagnetic compatibility – Requirements and tests .....	24
208      General requirements, tests and guidance for alarm systems in medical electrical equipment and medical electrical systems .....	28
Annexes .....	34
Annex AA (informative) Particular guidance and rationale .....	35
Annex BB (informative) Alarm diagrams of Clause 208/IEC 60601-1-8:2006 .....	43
Index of defined terms used in this particular standard.....	46
 Figure 201.101 – TRANSDUCER cable strain relief test.....	14
Figure 201.102 – Foam block test (see 201.11.1.2.2.104 and 201.11.1.2.2.105) .....	15
Figure 201.103 – Linearity and hysteresis test set-up – Gas mix chamber, assembled.....	20
Figure 201.104 – Linearity and hysteresis test set-up – Gas mix chamber, manufacturing dimensions .....	21
Figure 201.105 – Linearity and hysteresis test set-up – Gas mix chamber, dimensions of hose connector .....	22
Figure 202.101 – Set-up for radiated and conducted EMISSIONS testing according to 202.6.1.1.2 a) .....	25
Figure 202.102 – Set-up for radiated immunity test according to 202.6.2.3.2 .....	27
Figure BB.1 – NON-LATCHING ALARM SIGNALS without ALARM RESET .....	43
Figure BB.2 – NON-LATCHING ALARM SIGNALS with ALARM RESET .....	43
Figure BB.3 – LATCHING ALARM SIGNALS with ALARM RESET .....	44
Figure BB.4 – Two ALARM CONDITIONS with ALARM RESET.....	44

- - - - -

Table 201.101 – Distributed ESSENTIAL PERFORMANCE requirements .....	11
Table 201.102 – Required readings and tolerances .....	19
Table 201.103 – Calibration test gases .....	19
Table 208.101 – ALARM CONDITION priorities .....	29
Table 208.102 – Characteristics of the burst of auditory ALARM SIGNALS .....	30