



Designation: D1498 – 14 (Reapproved 2022)^{ε1}

Standard Test Method for Oxidation-Reduction Potential of Water¹

This standard is issued under the fixed designation D1498; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

^{ε1} NOTE—An editorial update was made to Section 1 in November 2022.

1. Scope

1.1 This test method covers the apparatus and procedure for the electrometric measurement of oxidation-reduction potential (ORP) in water. It does not deal with the manner in which the solutions are prepared, the theoretical interpretation of the oxidation-reduction potential, or the establishment of a standard oxidation-reduction potential for any given system. The test method described has been designed for the routine and process measurement of oxidation-reduction potential.

1.2 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.4 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 *ASTM Standards:*²

[D1129 Terminology Relating to Water](#)

[D1193 Specification for Reagent Water](#)

[D2777 Practice for Determination of Precision and Bias of Applicable Test Methods of Committee D19 on Water](#)

[D3370 Practices for Sampling Water from Flowing Process Streams](#)

3. Terminology

3.1 *Definitions*—For definitions of terms used in this test method, refer to Terminology [D1129](#).

3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *oxidation-reduction potential, n* —the electromotive force, E_m , developed between a noble metal electrode and a standard reference electrode.

3.2.1.1 *Discussion*—This oxidation-reduction potential (ORP) is related to the solution composition by:

$$E_m = E^o + 2.3 \frac{RT}{nF} \log[A_{ox}/A_{red}]$$

where:

E_m	=	ORP,
E^o	=	constant that depends on the choice of reference electrodes,
F	=	Faraday constant,
R	=	gas constant,
T	=	absolute temperature, °C + 273.15,
n	=	number of electrons involved in process reaction, and
A_{ox} and A_{red}	=	activities of the reactants in the process.

4. Summary of Test Method

4.1 This is a test method designed to measure the ORP which is defined as the electromotive force between a noble metal electrode and a reference electrode when immersed in a solution. The test method describes the equipment available to make the measurement, the standardization of the equipment and the procedure to measure ORP. The ORP electrodes are inert and measure the ratio of the activities of the oxidized to the reduced species present.

5. Significance and Use

5.1 Various applications include monitoring the chlorination/dechlorination process of water, recognition of oxidants/reductants present in wastewater, monitoring the cycle chemistry in power plants, and controlling the processing of cyanide and chrome waste in metal plating baths.

¹ This test method is under the jurisdiction of ASTM Committee [D19](#) on Water and is the direct responsibility of Subcommittee [D19.03](#) on Sampling Water and Water-Formed Deposits, Analysis of Water for Power Generation and Process Use, On-Line Water Analysis, and Surveillance of Water.

Current edition approved Nov. 1, 2022. Published December 2022. Originally approved in 1957. Last previous edition approved in 2014 as D1498 – 14. DOI: 10.1520/D1498-14R22E01.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

5.2 The measurement of ORP has been found to be useful in the evaluation of soils, for evaluating treatment design data at sites contaminated with certain chemicals, and in evaluating solid wastes.

6. Interferences

6.1 The ORP electrodes reliably measure ORP in nearly all aqueous solutions and in general are not subject to solution interference from color, turbidity, colloidal matter, and suspended matter.

6.2 The ORP of an aqueous solution is sensitive to change in temperature of the solution, but temperature correction is rarely done due to its minimal effect and complex reactions. Temperature corrections are usually applied only when it is desired to relate the ORP to the activity of an ion in the solutions.

6.3 The ORP of an aqueous solution is almost always sensitive to pH variations even to reactions that do not appear to involve hydrogen or hydroxyl ions. The ORP generally tends to increase with an increase in hydrogen ions and to decrease with an increase in hydroxyl ions during such reactions.

6.4 Reproducible oxidation-reduction potentials cannot be obtained for chemical systems that are not reversible. Most natural and ground waters do not contain reversible systems, or may contain systems that are shifted by the presence of air. The measurement of end point potential in oxidation-reduction titration is sometimes of this type.

6.5 If the metallic portion of the ORP electrode is sponge-like, materials absorbed from solutions may not be washed away, even by repeated rinsings. In such cases, the electrode may exhibit a memory effect, particularly if it is desired to detect a relatively low concentration of a particular species immediately after a measurement has been made in a relatively concentrated solution. A brightly polished metal electrode surface is required for accurate measurements.

6.6 The ORP resulting from interactions among several chemical systems present in mixed solutions may not be assignable to any single chemical.

7. Apparatus

7.1 *Meter*—Most laboratory pH meters can be used for measurements of ORP by substitution of an appropriate set of electrodes and meter scale. Readability to 1 mV is adequate. The choice will depend on the accuracy desired in the determination.

7.1.1 Most process pH meters can be used for measurement of ORP by substitution of an appropriate set of electrodes and meter scale. These instruments are generally much more rugged than those which are used for very accurate measurements in the laboratory. Usually, these more rugged instruments produce results that are somewhat less accurate and precise than those obtained from laboratory instruments. The choice of process ORP analyzer is generally based on how closely the characteristics of the analyzer match the requirements of the application. Typical factors which may be

considered include, for example, the types of signals which the analyzer can produce to drive external devices, and the span ranges available.

7.1.2 For remote ORP measurements the potential generated can be transmitted to an external indicating meter. Special shielded cable is required to transmit the signal.

7.2 *Reference Electrode*—A calomel, silver-silver chloride, or other reference electrode of constant potential shall be used. If a saturated calomel electrode is used, some potassium chloride crystals shall be contained in the saturated potassium chloride solution. If the reference electrode is of the flowing junction type, a slow outward flow of the reference-electrode solution is desired. To achieve this, the solution pressure inside the liquid junction should be somewhat in excess of that outside the junction. In nonpressurized applications this requirement can be met by maintaining the inside solution level higher than the outside solution level. If the reference electrode is of the nonflowing junction type, these outward flow and pressurization considerations shall not apply. The reference electrode and junction shall perform satisfactorily as required in the procedure for checking sensitivity described in 11.2.

7.3 *Oxidation-Reduction Electrode*—A noble metal is used in the construction of oxidation-reduction electrodes. The most common metals employed are platinum and gold; silver is rarely used. It is important to select a metal that is not attacked by the test solution. The construction of the electrode shall be such that only the noble metal comes in contact with the test solution. The area of the noble metal in contact with the test solution should be approximately 1 cm².

7.4 *Electrode Assembly*—A conventional electrode holder or support can be employed for laboratory measurements. Many different styles of electrode holders are suitable for various process applications such as measurements in an open tank, process pipe line, pressure vessel, or a high pressure sample line.

8. Reagents and Materials

8.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society.³ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

8.2 *Purity of Water*—References to water that is used for reagent preparation, rinsing, or dilution shall be understood to mean water that conforms to the quantitative specifications for type I or II reagent water of Specification D1193.

³ *ACS Reagent Chemicals, Specifications and Procedures for Reagents and Standard-Grade Reference Materials*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.