



Designation: E1061 – 01 (Reapproved 2022)

# Standard Specification for Direct-Reading Liquid Crystal Forehead Thermometers<sup>1</sup>

This standard is issued under the fixed designation E1061; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This specification covers direct-reading liquid crystal forehead thermometers designed to monitor trends or measure human forehead surface temperatures, or both.

1.2 Thermometers covered by this specification are generally available in continuous or intermittent reading modes. Thermometers meeting the requirements specified herein shall be considered as complying with this specification.

1.3 The following precautionary caveat pertains only to the test method portion, Section 5, of this specification. *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.4 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

## 2. Referenced Documents

2.1 *ASTM Standards:*<sup>2</sup>

E344 Terminology Relating to Thermometry and Hydrometry

## 3. Terminology

3.1 *Definitions*—For definitions of terms used in this specification, refer to Terminology E344.

3.2 *Definitions of Terms Specific to This Standard:*

3.2.1 *center green (CG) or mid green (MG), n*—that temperature which unifies the visual and instrumental evaluation methods and is defined by the equation:

$$CG = MG = \frac{SG + SB}{2} = T^{\circ}520$$

See 3.2.5.2 for description of T°520.

3.2.2 *color play, n*—the predictable sequence of colors exhibited by a liquid crystal formulation as it passes through its active temperature range. For example, as temperature increases, a formulation exhibits successive tan, red, green, and blue colors.

3.2.3 *manufacturing lot, n*—(1) in the case of continuous manufacturing processes, a *lot* is a specifically identified amount produced in a unit of time or quantity in a manner that assumes its having uniform characteristics and quality within specified limits. (2) In the case of batch processes, a *lot* means a batch or specifically identified portion of a batch assumed to have uniform characteristics and quality within specified limits.

3.2.4 *sequencing, n*—a characteristic whereby the thermal profiles of the liquid crystal formulations of a given thermometer follow each other in an orderly predetermined manner.

3.2.5 *specific color phenomena, n* (using instrumental methods of evaluation):

3.2.5.1 *blue*—that temperature at which the intensity of 470 nm light reflected by the liquid crystal is maximum, symbolized as T°470.

3.2.5.2 *green*—that temperature at which the intensity of 520 nm light reflected by the liquid crystal is maximum, symbolized as T°520.

3.2.5.3 *red*—that temperature at which the intensity of 650 nm light reflected by the liquid crystal is maximum, symbolized as T°650.

3.2.6 *start of blue (SB), n*—that temperature at which the liquid crystal first begins to reflect blue light, which is defined as light having a wavelength of 491 nm.

3.2.7 *start of green (SG), n*—that temperature at which the liquid crystal first begins to reflect green light, which is defined as light having a wavelength of 575 nm.

3.2.8 *start of red (SR), n*—that temperature at which the liquid crystal first begins to reflect red light, which is defined as light having a wavelength of 675 nm.

3.2.9 *thermal profile, n*—temperatures at which specific color phenomena occur in a liquid crystal thermometer.

<sup>1</sup> This specification is under the jurisdiction of ASTM Committee F04 on Medical and Surgical Materials and Devices and is the direct responsibility of Subcommittee F04.33 on Medical/Surgical Instruments.

Current edition approved Feb. 1, 2022. Published February 2022. Originally approved in 1985. Last previous edition approved in 2014 as E1061 – 01 (2014). DOI: 10.1520/E1061-01R22.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.