



Designation: F3224 – 17

Standard Test Method for Evaluating Growth of Engineered Cartilage Tissue using Magnetic Resonance Imaging¹

This standard is issued under the fixed designation F3224; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This standard is intended as a standard test method for engineered cartilage tissue growth evaluation using MRI.

1.2 This standard is intended for use in the development of tissue engineering regenerative medical products for cartilage damages, such as in knee, hip, or shoulder joints.

1.3 This standard has been prepared for evaluation of engineered cartilage tissue growth at the preclinical stage and summarizes results from tissue growth evaluation of tissue-engineered cartilage in a few notable cases using water spin-spin relaxation time, T_2 , *in vitro* and *in vivo* in small animal models.

1.4 This standard uses the change in mean T_2 values as a function of growth time to evaluate the tissue growth of engineered cartilage.

1.5 This standard provides a method to remove the scaffold contribution to the tissue growth evaluation.

1.6 Information in this standard is intended to be applicable to most porous natural and synthetic polymers used as a scaffold in engineered cartilage, such as alginate, agarose, collagen, chitosan, and poly-lactic-co-glycolic acid (PLGA). However, some materials (both synthetic and natural) may require unique or varied methods of MRI evaluation that are not covered in this test method.

1.7 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

1.8 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

¹ This test method is under the jurisdiction of ASTM Committee F04 on Medical and Surgical Materials and Devices and is the direct responsibility of Subcommittee F04.44 on Assessment for TEMPs.

Current edition approved Nov. 1, 2017. Published February 2018. DOI: 10.1520/F3224-17.

2. Referenced Documents

2.1 The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document applies.

2.2 ASTM Standards:²

F2312 [Terminology Relating to Tissue Engineered Medical Products](#)

F2529 [Guide for *in vivo* Evaluation of Osteoinductive Potential for Materials Containing Demineralized Bone \(DBM\)](#)

F2603 [Guide for Interpreting Images of Polymeric Tissue Scaffolds](#)

F2664 [Guide for Assessing the Attachment of Cells to Biomaterial Surfaces by Physical Methods](#)

F2978 [Guide to Optimize Scan Sequences for Clinical Diagnostic Evaluation of Metal-on-Metal Hip Arthroplasty Devices using Magnetic Resonance Imaging](#)

2.3 ISO Standard:³

ISO/TR 16379-2014 [Tissue-engineered medical products — Evaluation of anisotropic structure of articular cartilage using DT \(Diffusion Tensor\)-MR Imaging](#)

3. Terminology

3.1 Definitions of Terms Specific to This Standard:

3.1.1 *biomaterial, n*—any substance (other than a drug), synthetic or natural, that can be used as a system or part of a system that treats, augments, or replaces any tissue, organ, or function of the body. **F2664**

3.1.2 *chondrocyte, n*—a cell that has secreted the matrix of cartilage and becomes embedded in it.

3.1.3 *chondrogenic differentiation, n*—the biological process of stem cells changing their lineage into chondrocytes. If the starting cells are chondrocytes, this term refers to differentiation of cells into the same phenotype.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Available from American National Standards Institute (ANSI), 25 W. 43rd St., 4th Floor, New York, NY 10036, <http://www.ansi.org>.

3.1.4 *chondrogenic extracellular matrix (chondrogenic ECM)*, *n*—an extracellular matrix containing cartilaginous matrix proteins such as proteoglycan, collagen type II, collagen type X and other matrix proteins found in cartilage.

3.1.5 *echo time (TE)*, *n*—time after 90° pulse in an MRI pulse sequence until an echo signal is formed.

3.1.6 *fast low angle shot (FLASH) MRI*, *n*—a gradient echo MRI acquisition technique with low flip angle radiofrequency pulse excitation and short repetition time for fast image acquisition.

3.1.7 *field of view (FOV)*, *n*—MR image acquisition parameter that defines the dimensions of the imaging plane (expressed in cm × cm or mm × mm).

3.1.8 *histological assessment of engineered cartilage tissue growth*, *n*—histological assessment is used to assess the presence of cartilage extracellular matrix proteins in the engineered cartilage to evaluate the tissue growth (e.g. Safranin O staining for proteoglycan assessment).

3.1.9 *hydrogel*, *n*—a water-based open network of polymer chains that are cross-linked either chemically or through crystalline junctions or by specific ionic interactions. **F2603**

3.1.10 *in-plane resolution*, *n*—the spatial resolution of an image (typically expressed in mm × mm or μm × μm). It is given by = FOV/acquired matrix size.

3.1.11 *magnetic resonance imaging (MRI)*, *n*—an imaging technique that uses static and time-varying magnetic fields to provide tomographic images of tissue by the magnetic resonance of nuclei. **F2978**

3.1.12 *matrix size*, *n*—the number of pixels in each image dimension of FOV.

3.1.13 *mesenchymal stem cell (MSC)*, *n*—a multipotent cell derived from mesenchyme that is capable of proliferating and differentiating in chondrogenic lineage and can produce a cartilage extracellular matrix.

3.1.14 *multi slice multi echo (MSME) MRI*, *n*—an MRI pulse sequence for the measurement of T₂ where a series of 180° RF pulses (number of echoes) is followed by a 90° RF pulse in a multi-slice MRI pulse sequence. This pulse sequence is the MRI extension of similar nuclear magnetic resonance (NMR) spectroscopy sequence named Carr-Purcell-Meiboom-Gill (CPMG) echo train pulse sequence for T₂ measurement.

3.1.15 *number of averages (NA)*, *n*—the number of times an identical MRI experiment is repeated to improve the SNR.

3.1.16 *pulse sequence*, *n*—programmed train of RF and gradient pulses. In MRI, it is a time protocol for obtaining images.

3.1.17 *quantitative real-time polymerase chain reaction (qRPCR)*, *n*—a laboratory technique for the detection, selection, and amplification of specific gene transcripts based on their genetic sequence. Commonly, it is used to assess the presence of chondrogenic markers such as Sox9, RUNX2, ECM proteins, etc. in a tissue-engineered cartilage.

3.1.18 *radiofrequency pulse (RF pulse)*, *n*—a short duration radiofrequency electromagnetic pulse used for changing the direction of magnetization vector.

3.1.19 *rapid acquisition with refocused echoes (RARE) MRI*, *n*—an MRI pulse sequence for fast image acquisition. This MRI pulse sequence is characterized by a series of 180° RF rephasing pulses followed by a 90° RF pulse, with each echo is individually phase-encoded for fast image acquisition.

3.1.20 *region of interest (ROI)*, *n*—a user-defined area of an image in which parameter of interest is calculated.

3.1.21 *relaxation rate (R₂)*, *n*—inverse of spin-spin relaxation time (R₂ = 1/T₂).

3.1.22 *repetition time (TR)*, *n*—time interval between consecutive 90° RF pulses or the time interval when the basic unit of MRI pulse sequence is repeated. **ISO/TR 16379-2014**

3.1.23 *scaffold*, *n*—three-dimensional natural or synthetic biomaterial typically made out of one or more polymers (natural or synthetic) and used as a skeleton for cell seeding. **F2603**

3.1.24 *signal to noise ratio (SNR)*, *n*—the ratio of the amplitude of any signal of interest to the amplitude of the average background noise which includes both coherent and non-coherent types of noise.

3.1.25 *slice thickness*, *n*—the thickness of the 2D imaging plane in an MRI image. **ISO/TR 16379-2014**

3.1.26 *spin echo (SE) MRI*, *n*—a method for acquiring MR images based on the spin-echo pulse sequence.

3.1.27 *spin-spin relaxation time (T₂)*, *n*—T₂ refers to the characteristic exponential time constant of the transverse magnetization. This is typically the time taken for the transverse magnetization to decrease to 37% of the initial value. It is typically depicted in milliseconds (ms).

3.1.28 *stem cell*, *n*—an undifferentiated cell that is capable of developing into many different cell types.

3.1.29 *voxel*, *n*—the minimum unit volume of a three-dimensional MRI image. **ISO/TR 16379-2014**

4. Significance and Use

4.1 Tissue-engineered cartilage is prepared by seeding stem cells or chondrocytes in a three-dimensional biodegradable scaffold under controlled growth conditions. It is expected that the cells will differentiate towards chondrogenic lineage and produce an ample amount of cartilage extracellular matrix proteins, proteoglycans, and collagen type-II. Longitudinal assessment is needed weekly for the first few weeks *in vitro* and monthly at a later stage *in vivo* to determine the growth rate of tissue-engineered cartilage. Traditional testing methods such as histological staining, mechanical testing, and qPCR are invasive, destructive, and cannot be performed *in vivo* after the transplantation of engineered tissue as a regenerative treatment. In the regenerative medicine of cartilage, it is important to evaluate whether the implanted tissue regenerates as an articular cartilage over time. MRI is the only available non-invasive imaging modality that is utilized for post-operative monitoring and assessment of cartilage regeneration in clinics. Therefore, it is important to evaluate tissue-engineered cartilage using MRI at the preclinical stage as well.

4.2 Preclinical *in vivo* assessment of tissue-engineered cartilage is performed in small animal models such as mice, rats




FIG. 1 The change in the water relaxation time T_2 as a function of the magnetic field and the tumbling rate of the water molecule using BPP theory of relaxation (1). Note that the tumbling rate of the water molecule decreases with increasing tissue growth. The blue arrow shows the direction of change of the relaxation time, T_2 , as a function of the tissue growth.

or rabbits, and in large animal models such as goats, pigs, and horses. It is possible to evaluate engineered cartilage tissue growth at each stage of development non-invasively using MRI. This may reduce the number of animals needed for the assessment and will provide a good estimate of cartilage regeneration.

4.3 Parametric MRI technique allows non-invasive quantitative assessment of tissue growth *in vitro* and *in vivo*. When the amount of extracellular matrix increases over time, the interaction of the water molecule with its surroundings changes, and this creates a change in T_2 . The amount of change in T_2 is directly correlated with the amount of matrix generated with high sensitivity and specificity. The T_2 MRI is thus used to observe tissue growth for use commonly in longitudinal diagnosis following cell seeding in a scaffold *in vitro* or following tissue implantation *in vivo*.

4.4 The T_2 MRI for preclinical evaluation of engineered cartilage takes into account the presence of a scaffold in the developing tissue-engineered cartilage. These data are published in refereed journals and book chapters, and included here as a guide for preclinical quantitative evaluation for engineered cartilage tissue growth (2-12).⁴ Additional data utilizing T_2 MRI for tissue growth evaluation of engineered cartilage can be found in the references (13-15).

⁴ The boldface numbers in parentheses refer to the list of references at the end of this standard.

4.5 Magnetic resonance parameters of water protons in tissue are sensitive to the tissue microstructure. In cartilage tissue engineering, cells produce primarily cartilage extracellular matrix proteins, proteoglycans, and collagen, type-II. As tissue matures with the production of ECM, the matrix changes the environment around water molecules. The water nuclear spins find several new pathways for relaxation and the T_2 generally is lower from the original value. Fig. 1 shows the effect of water tumbling rate and magnetic field strength on T_2 . As shown by the blue arrow, when the engineered cartilage tissue matures, the tumbling rate of the water molecule is lower and as a result, the T_2 is lower. The reduction of T_2 as a function of tissue growth is the basis of engineered cartilage assessment using MRI. Fig. 1 also shows that this principle holds true from low to high magnetic field strengths (1.5 T – 11.7 T) that are commonly used in MRI assessment.

4.6 As shown in Fig. 1, the change in T_2 is dependent on the magnetic field strength and initial tumbling rate of the water molecule that signifies its surrounding.

4.7 The principle of reduced T_2 with increased tissue growth generally holds true for scaffold-free cartilage tissue engineering. However, in scaffold-based cartilage tissue engineering, the following relationship should be used to assess the tissue growth (3, 6):

$$R_2(ECM) = R_2(TEC) - R_2(Control) \quad (1)$$