



Designation: F3294 – 18

Standard Guide for Performing Quantitative Fluorescence Intensity Measurements in Cell-based Assays with Widefield Epifluorescence Microscopy¹

This standard is issued under the fixed designation F3294; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This guidance document has been developed to facilitate the collection of microscopy images with an epifluorescence microscope that allow quantitative fluorescence measurements to be extracted from the images. The document is tailored to cell biologists that often use fluorescent staining techniques to visualize components of a cell-based experimental system. Quantitative comparison of the intensity data available in these images is only possible if the images are quantitative based on sound experimental design and appropriate operation of the digital array detector, such as a charge coupled device (CCD) or a scientific complementary metal oxide semiconductor (sCMOS) or similar camera. Issues involving the array detector and controller software settings including collection of dark count images to estimate the offset, flat-field correction, background correction, benchmarking of the excitation lamp and the fluorescent collection optics are considered.

1.2 This document is developed around epifluorescence microscopy, but it is likely that many of the issues discussed here are applicable to quantitative imaging in other fluorescence microscopy systems such as fluorescence confocal microscopy. This guide is developed around single-color fluorescence microscopy imaging or multi-color imaging where the measured fluorescence is spectrally well separated.

1.3 Fluorescence intensity is a relative measurement and does not in itself have an associated SI unit. This document does discuss metrology issues related to relative measurements and experimental designs that may be required to ensure quantitative fluorescence measurements are comparable after changing microscope, sample, and lamp configurations.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety, health, and environmental practices and determine the applicability of regulatory limitations prior to use.*

¹ This guide is under the jurisdiction of ASTM Committee F04 on Medical and Surgical Materials and Devices and is the direct responsibility of Subcommittee F04.46 on Cell Signaling.

Current edition approved Oct. 1, 2018. Published October 2018. DOI: 10.1520/F3294-18.

1.5 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:²

[E131 Terminology Relating to Molecular Spectroscopy](#)

[E284 Terminology of Appearance](#)

[E2186 Guide for Determining DNA Single-Strand Damage in Eukaryotic Cells Using the Comet Assay](#)

[E2642 Terminology for Scientific Charge-Coupled Device \(CCD\) Detectors](#)

[E2719 Guide for Fluorescence—Instrument Calibration and Qualification](#)

[E2825 Guide for Forensic Digital Image Processing](#)

[F2944 Test Method for Automated Colony Forming Unit \(CFU\) Assays—Image Acquisition and Analysis Method for Enumerating and Characterizing Cells and Colonies in Culture](#)

[F2997 Practice for Quantification of Calcium Deposits in Osteogenic Culture of Progenitor Cells Using Fluorescent Image Analysis](#)

[F2998 Guide for Using Fluorescence Microscopy to Quantify the Spread Area of Fixed Cells](#)

2.2 ISO Standards:³

[ISO 13653 Measurement of relative irradiance in the image field](#)

[ISO/IEC 10918-1:1994 Digital compression and coding of continuous-tone still images: Requirements and guidelines](#)

[ISO/TR 12033:2009 Guidance for the selection of document image compression methods](#)

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Available from American National Standards Institute (ANSI), 25 W. 43rd St., 4th Floor, New York, NY 10036, <http://www.ansi.org>.

2.3 Other Documents:

- SWGDE/SWGIT Glossary SWGDE and SWGIT Digital & Multimedia Evidence Glossary, updated June 8, 2012⁴
 U.S. Food and Drug Administration (FDA) Guidance, Technical Performance Assessment of Digital Pathology Whole Slide Imaging Devices⁵
 European Machine Vision Association (EMVA) Standard 1288 Standard for Characterization and Presentation of Specification Data for Image Sensors and Cameras⁶

3. Terminology

3.1 Definitions of Terms Specific to This Standard:

3.1.1 *CCD bias, n*—the minimum analog offset added to the signal before the A/D converter to ensure a positive digital output each time a signal is read out. **E2642**

3.1.2 *charge-coupled device (CCD), n*—a silicon-based semiconductor chip consisting of a two-dimensional matrix of photo sensors or pixels. **E2642**

3.1.3 *complementary metal oxide semiconductor (CMOS), n*—technology widely used to manufacture electronic devices and image sensors similar to CCDs. In a CMOS sensor, each pixel has its own charge-to-voltage conversion circuit, and the sensor often also includes amplifiers, noise correction, and digitization circuits. Due to the additional components associated with each pixel, the sensitivity to light is lower than with a CCD, the signal is noisier, and the uniformity is lower. But the sensor can be built to require less off-chip circuitry for basic operation. **E2642**

3.1.4 *dynamic range, n*—the ratio of the full well saturation charge to the system noise level. It represents the ratio of the brightest and darkest signals a detector can measure in a single measurement. **E2642**

3.1.5 *electron-multiplying CCD (EMCCD), n*—type of CCD that has a two-way readout register, that is, the shift register and the gain register, each with its own output amplifier. When the charge is read out through the shift register, the detector works like a standard CCD detector, and when the charge is read out through the gain register, it undergoes charge amplification as a result of a different electrode structure embedded underneath the pixels of this register. **E2642**

3.1.6 *exposure time, n*—the length of time for which a CCD accumulated charge. **E2642**

3.1.7 *fluorescence, n*—the emission of radiant energy from an atom, molecule, or ion resulting from absorption of a photon and a subsequent transition to the ground state without a change in total spin quantum number. **E131**

3.1.8 *frame, n*—one full image that is read out of a CCD. **E2642**

3.1.9 *full well capacity, n*—the maximum number of photoelectrons that can be collected on a single pixel in the image area or in the horizontal register of a CCD. It is typically specified in terms of number of electrons. **E2642**

3.1.10 *irradiance, E, E_e, n*—the radiant flux incident per unit area. **E284**

3.1.11 *lossless compression, n*—compression in which no data are lost and all data can be retrieved in their original form. **SWGDE/SWGIT Glossary**

3.1.12 *lossy compression, n*—compression in which data are lost and cannot be retrieved in their original form. **SWGDE/SWGIT Glossary**

3.1.13 *photobleaching, n*—loss of emission or absorption intensity by a sample as a result of exposure to optical radiation. **E2719**

3.1.14 *pixel, n*—abbreviation for picture element. The smallest unit in an optical device in which charge is collected as a signal. CCD detectors typically have 26 μm square pixels; however, pixel sizes of 8, 13, 16, and 20 μm square are also available. **E2642**

3.1.15 *radiant energy, n*—energy transmitted as electromagnetic radiation. **E284**

3.1.16 *radiant flux, Φ, n*—the time rate of flow of radiant energy; radiant power. **E284**

3.1.17 *region of interest (ROI), n*—user-defined portion of the image area in which data will be acquired. The remainder of the image area will be discarded. **E2642**

4. Summary of Guide

4.1 Wide-field fluorescence microscopy is an optical imaging technique that relies on illumination of the entire field of view of a fluorescence microscope and simultaneous detection of the emitted fluorescence from all or a sub-region of the field of view using a camera. The emitted fluorescence can be measured as an intensity value in fluorescence microscopy, which is computed by summing together the intensity values from a group of individual pixels in a digital image acquired using a digital camera, such as a CCD, sCMOS, or EMCCD. A relative intensity measurement (RIM) is determined as the ratio of one intensity measurement to another, the result of which should be an accurate estimate of the ratio of the irradiance from part or all of a specimen to the irradiance from part or all of the same or another specimen.

4.2 The quantitative comparison of RIMs can be compromised or invalidated by many possible factors including the non-uniformity of intensities across the field of view of the microscope, the presence of an offset in the pixel values in the recorded digital image, the intensity signals in the image exceeding the linear dynamic range of the camera, or the inaccurate recording of the pixel values in image data files due to factors such as a lossy compression operation or unexpected modification of the pixel bit depth when saving each file. The quantitative comparison of RIMs can also be compromised by low signal-to-noise ratio of the measured light intensities or instability in the optical power of the illumination source.

⁴ Available from Scientific Working Group on Imaging Technology (SWGIT), <http://www.swgit.org>

⁵ Available from U.S. Department of Health and Human Services, Food and Drug Administration (FDA), Center for Devices and Radiological Health (CDRH). <https://www.fda.gov/downloads/MedicalDevices/DeviceRegulationandGuidance/GuidanceDocuments/UCM435355.pdf>. You may also send an e-mail request to CDRH-Guidance@fda.hhs.gov to receive a copy of the guidance. Please use the document number 1400053 to identify the guidance you are requesting.

⁶ Available from European Machine Vision Association. <http://www.emva.org/standards-technology/emva-1288/>.

4.3 This guide provides a list of corrections and normalizations that are required so that RIMs can be accurately compared within and between images. This guidance document also includes a list of references to peer-reviewed, published methods that can aid in the qualification of instrument performance in [Appendix X1](#). The collection of these resources should be useful in the design of robust cell-based assays that used quantitative fluorescence microscopy for data collection.

5. Significance and Use

5.1 *Overview of Measurement System*—Relative intensity measurements made by widefield epifluorescence microscopy are used as part of cell-based assays to quantify attributes such as the abundance of probe molecules (see ASTM [F2997](#)), fluorescently labeled antibodies, or fluorescence protein reporter molecules. The general procedure for quantifying relative intensities is to acquire digital images, then to perform image analysis to segment objects and compute intensities. The raw digital images acquired by epifluorescence microscopy are not typically amenable to relative intensity quantification because of the factors listed in [4.2](#). This guide offers a checklist of potential sources of bias that are often present in fluorescent microscopy images and suggests approaches for storing and normalizing raw image data to assure that computations are unbiased.

5.2 *Areas of Application*—Widefield fluorescence microscopy is frequently used to measure the location and abundance of fluorescent probe molecules within or between cells. In instances where RIM comparisons are made between a region of interest (ROI) and another ROI, accurate normalization procedures are essential to the measurement process to minimize biased results. Example use cases where this guidance document may be applicable include:

5.2.1 Characterization of cell cycle distribution by quantifying the abundance of DNA in individual cells ([1](#)).⁷

5.2.2 Measuring the area of positively stained mineralized deposits in cell cultures (ASTM [F2997](#)).

5.2.3 Quantifying the spread area of fixed cells (ASTM [F2998](#)).

5.2.4 Determining DNA damage in eukaryotic cells using the comet assay (ASTM [E2186](#)).

5.2.5 The quantitation of a secondary fluorescent marker that provides information related to the genotype, phenotype, biological activity, or biochemical features of a colony or cell (ASTM [F2944](#)).

6. Measurement Bias

6.1 Sources of bias in relative intensity measurements are listed below:

6.1.1 *CCD Bias*—The detectors used in widefield fluorescence microscopy are typically scientific complementary metal oxide semiconductor (sCMOS) sensors, charge-coupled devices (CCDs), electron multiplying (EM)-CCDs, or similar types of arrayed photodetectors ([2](#)). Regardless of the detector, the recorded digital signal in the absence of incident light,

called the CCD bias (a.k.a. bias current offset or dark counts), is often not zero. The CCD bias is an offset that is added to each pixel in the digital image. Accurate determination of the value of the CCD bias is critical as it will need to be subtracted from the raw image.

6.1.1.1 A CCD bias that is less than zero is problematic. Digital images are typically saved in files that store only positive integer values. If the CCD bias is less than zero, an unknown offset is subtracted from each pixel value and relative intensity comparisons will not be possible.

6.1.2 *Linear Dynamic Range*—Images of the specimen under evaluation must be collected with the signal within the linear dynamic range of the detector. Signals that are at or below the noise floor of the camera will not be detected. Similarly, signals that are above the detector saturation are no longer in the linear range of the camera and cannot be used in relative intensity evaluations.

6.1.3 *Non-Uniform Field*—The intensities measured from a uniformly fluorescent sample are typically not uniform across the field of view. Field non-uniformities can arise from many factors, including non-uniform illumination, vignetting, and non-uniformities in the detector (ISO 13653 and ([3](#), [4](#))). This means that the measured fluorescence intensity is dependent on its position within the field of view. If measurements are to be made in a region of the field of view with uneven illumination, a flatfield correction should be applied. An appropriate flatfield correction will result in measured intensities that are not dependent on their location within the field of view and can be compared.

6.1.4 *Save Raw Images or Use Lossless Compression*—Many of the software packages that are used for controlling image capture from digital cameras offer the opportunity to save images in a lossy compression format (e.g. jpeg). This form of compression can alter the intensity data in a non-linear fashion, leading to unpredictable biases in relative intensity measurements. It is best to save raw, non-compressed image data or a lossless compression format (e.g. tiff) for images that are intended to be used to make relative intensity measurements. Additional modifications to the pixel values of an image can occur due to unexpected settings on the image analysis software. For example, bit truncation or bit depth conversion can occur on a saved image. It is worthwhile to evaluate the histogram of pixel intensities on the saved image and compare it to the histogram of the collected image to ensure the image is appropriately saved. A good resource for information on lossless and lossy image compression formats can be found in ASTM [E2825-12](#), ISO/TR 12033:2009, and ISO/IEC 10918-1:1994.

7. Normalization Strategies for Sources of Bias

7.1 *CCD Bias*—To measure the CCD bias, a dark frame must be collected in the absence of illumination incident on the detector. This estimate for the CCD bias will only be accurate for the detector settings for which the image was taken. If any detector settings are changed, such as the temperature, gain, or binning, the CCD bias estimate may no longer be valid and may need to be remeasured. The CCD bias may change over the course of a data acquisition run caused, for example, by the

⁷ The boldface numbers in parentheses refer to the list of references at the end of this standard.