

INTERNATIONAL STANDARD

ISO
9029

First edition
1990-12-15

Crude petroleum — Determination of water — Distillation method

*Pétrole brut — Détermination de la teneur en eau — Méthode de
distillation*



Reference number
ISO 9029:1990(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

International Standard ISO 9029 was prepared by Technical Committee ISO/TC 28, *Petroleum products and lubricants*.

Annex A forms an integral part of this International Standard.

© ISO 1990

All rights reserved. No part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case Postale 56 • CH-1211 Genève 20 • Switzerland

Printed in Switzerland

Crude petroleum — Determination of water — Distillation method

WARNING — The use of this International Standard may involve hazardous materials, operations and equipment. This standard does not purport to address all of the safety problems associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

1 Scope

This International Standard specifies a method for determining water in crude oil by distillation. The precision data have only been determined for water contents up to 1 % (V/V).

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 383:1976, *Laboratory glassware — Interchangeable conical ground joints*.

ISO 3170:1988, *Petroleum liquids — Manual sampling*.

ISO 3171:1988, *Petroleum liquids — Automatic pipeline sampling*.

ISO 4259:1979, *Petroleum products — Determination and application of precision data in relation to methods of test*.

ISO 5280:1979, *Xylene for industrial use — Specification*.

3 Significance

A knowledge of the water content of crude oil is important in the refining, purchase, sale and transfer of products.

The amount of water as determined by this method is used to correct the volume involved in the custody transfer of oil.

4 Principle

A test portion is heated under reflux conditions with a water-immiscible solvent which co-distills with the water in the sample. Condensed solvent and water are continuously separated in a trap. The water settles in the graduated section of the trap, and the solvent returns to the distillation flask.

5 Apparatus

Usual laboratory apparatus, together with the following:

5.1 General.

The recommended apparatus, shown in figure 1, consists of a glass distillation flask, a condenser, a graduated glass trap and a heater. Other types of apparatus may be used for this International Standard, provided it can be demonstrated that they operate within the precision established, in accordance with ISO 4259, with the preferred apparatus.