

---

---

**Anaesthetic and respiratory  
equipment — Passive humidifiers**

*Matériel d'anesthésie et de réanimation respiratoire —  
Humidificateurs passifs*





**COPYRIGHT PROTECTED DOCUMENT**

© ISO 2018

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office  
CP 401 • Ch. de Blandonnet 8  
CH-1214 Vernier, Geneva  
Phone: +41 22 749 01 11  
Fax: +41 22 749 09 47  
Email: [copyright@iso.org](mailto:copyright@iso.org)  
Website: [www.iso.org](http://www.iso.org)

Published in Switzerland

# Contents

	Page
<b>Foreword</b> .....	<b>v</b>
<b>Introduction</b> .....	<b>vi</b>
<b>1 * Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>2</b>
<b>3 Terms and definitions</b> .....	<b>3</b>
<b>4 General requirements for testing</b> .....	<b>5</b>
4.1 Water level.....	5
4.2 PASSIVE HUMIDIFIER test conditions.....	6
4.3 * Gas flowrate and leakage specifications.....	6
4.4 * PASSIVE HUMIDIFIER testing errors.....	6
<b>5 General requirements</b> .....	<b>6</b>
5.1 Mechanical BASIC SAFETY for all PASSIVE HUMIDIFIERS.....	6
5.1.1 General.....	6
5.1.2 * Requirements for instability from unwanted lateral movement.....	6
5.1.3 * Requirements for audible acoustic energy.....	7
5.1.4 * Overflow.....	8
5.1.5 * Overpressure requirement.....	8
5.2 Compatibility requirement.....	9
5.3 General requirements for mechanical strength.....	9
<b>6 Identification, marking and ACCOMPANYING DOCUMENTATION</b> .....	<b>10</b>
6.1 Legibility and durability of markings.....	10
6.1.1 Legibility.....	10
6.1.2 Durability.....	10
6.2 Markings on the outside of the PASSIVE HUMIDIFIER or its parts.....	10
6.2.1 Identification.....	10
6.2.2 Additional requirements.....	11
6.2.3 Requirements for physiological effects.....	12
6.2.4 Requirements for packaging.....	12
6.2.5 Symbols.....	12
6.3 Units of measurement.....	12
6.4 Instructions for use.....	12
6.4.1 Identification.....	12
6.4.2 General requirements.....	13
6.4.3 * Requirements for warnings and safety notices.....	14
6.4.4 Requirements for installation.....	14
6.4.5 Requirements for start-up PROCEDURE.....	14
6.4.6 * Requirements for operating instructions.....	14
6.4.7 Requirements for CLEANING, DISINFECTION, and STERILIZATION.....	16
6.4.8 Requirements for maintenance.....	16
6.4.9 Requirements for ACCESSORIES, supplementary equipment and used material.....	16
6.4.10 Unique version identifier.....	16
6.5 Technical description.....	17
<b>7 * HUMIDIFICATION OUTPUT</b> .....	<b>17</b>
<b>8 Systems requirements</b> .....	<b>17</b>
<b>9 Specific SINGLE FAULT CONDITIONS</b> .....	<b>18</b>
<b>10 * CLEANING and DISINFECTION</b> .....	<b>18</b>
10.1 General.....	18
10.2 Home healthcare environment.....	18
<b>11 * BREATHING SYSTEM connectors and ports</b> .....	<b>18</b>
11.1 General.....	18