
**Sterilization of health care products —
Low temperature vaporized hydrogen
peroxide — Requirements for the
development, validation and routine
control of a sterilization process for
medical devices**

*Stérilisation des produits de santé — Vapeur de peroxyde d'hydrogène
à basse température — Exigences pour la mise au point, la validation
et le contrôle de routine d'un procédé de stérilisation pour dispositifs
médicaux*





COPYRIGHT PROTECTED DOCUMENT

© ISO 2022

All rights reserved. Unless otherwise specified, or required in the context of its implementation, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
CP 401 • Ch. de Blandonnet 8
CH-1214 Vernier, Geneva
Phone: +41 22 749 01 11
Email: copyright@iso.org
Website: www.iso.org

Published in Switzerland

Contents

	Page
Foreword.....	v
Introduction.....	vi
1 Scope.....	1
1.1 Inclusions.....	1
1.2 Exclusions.....	1
2 Normative references.....	2
3 Terms and definitions.....	2
4 Quality management system elements.....	10
5 Sterilizing agent characterization.....	11
5.1 General.....	11
5.2 Sterilant and sterilizing agent.....	11
5.3 Microbicidal effectiveness.....	11
5.4 Effects on materials.....	12
5.5 Safety and the environment.....	12
6 Process and equipment characterization.....	13
6.1 General.....	13
6.2 Process characterization.....	13
6.3 Equipment characterization.....	14
7 Product definition.....	15
8 Process definition.....	16
9 Validation.....	17
9.1 General.....	17
9.2 Installation qualification (IQ).....	18
9.2.1 General.....	18
9.2.2 Equipment.....	18
9.3 Operational qualification (OQ).....	19
9.4 Performance qualification (PQ).....	19
9.5 Review and approval of validation.....	21
10 Routine monitoring and control.....	21
11 Product release from sterilization.....	22
12 Maintaining process effectiveness.....	22
12.1 General.....	22
12.2 Recalibration.....	22
12.3 Maintenance of equipment.....	23
12.4 Requalification.....	23
12.5 Assessment of change.....	23
Annex A (normative) Factors to be considered in selection of microorganisms for demonstrating microbicidal effectiveness.....	24
Annex B (normative) Approach 1 — Process definition based on inactivation of the microbial population in its natural state (bioburden method).....	26
Annex C (normative) Approach 2 — Process definition based on inactivation of reference microorganisms and knowledge of bioburden (BI/bioburden method).....	28
Annex D (normative) Approach 3 — Conservative process definition based on inactivation of reference microorganisms (overkill method).....	30
Annex E (informative) Guidance on application of this document.....	33
Annex F (informative) Schematic example of a VH2O2 sterilization cycle.....	56