



# AEROSPACE RECOMMENDED PRACTICE

ARP598™

REV. D

Issued 1960-03  
Revised 2003-01  
Cancelled 2016-06

Superseded by AS598

Aerospace Microscopic Sizing and Counting of Particulate Contamination for Fluid Power Systems

## RATIONALE

This document is Cancelled and superseded by AS598.

## CANCELLATION NOTICE

This Technical Report has been declared "CANCELLED" as of June 2016 and has been superseded by AS598. By this action, this document will remain listed in the respective index, if applicable. Cancelled Technical Reports are available from SAE.

SAE Technical Standards Board Rules provide that: "This report is published by SAE to advance the state of technical and engineering sciences. The use of this report is entirely voluntary, and its applicability and suitability for any particular use, including any patent infringement arising therefrom, is the sole responsibility of the user."

SAE reviews each technical report at least every five years at which time it may be revised, reaffirmed, stabilized, or cancelled. SAE invites your written comments and suggestions.

Copyright © 2016 SAE International

All rights reserved. No part of this publication may be reproduced, stored in a retrieval system or transmitted, in any form or by any means, electronic, mechanical, photocopying, recording, or otherwise, without the prior written permission of SAE.

TO PLACE A DOCUMENT ORDER: Tel: 877-606-7323 (inside USA and Canada)

Tel: +1 724-776-4970 (outside USA)

Fax: 724-776-0790

Email: CustomerService@sae.org

SAE WEB ADDRESS:

<http://www.sae.org>

SAE values your input. To provide feedback  
on this Technical Report, please visit  
<http://www.sae.org/technical/standards/ARP598D>

## FOREWORD

Of the methods available for determining the level of particulate contamination in aerospace hydraulic fluids the most predominant have been gravimetric, microscopic counting (this document), automatic particle counting and membrane.

Gravimetric analysis results cannot be related to any specific particle count or related contamination class. Automatic Particle Counters (APCs) calibrated with AC fine test dust provide results classes compatible with this document; however AC test dust is no longer manufactured and use of this calibration procedure is being phased out. Membrane standards can be made compatible to both the above. Results obtained by each of these three methods from identical bottled samples may not be equivalent due to individual idiosyncrasies in each technique. Other APC variants, which will be coming into increased use, are calibrations using new ISO test dusts, on-line counting systems and portable APCs. Current results indicate these will not be compatible with this document.

## TABLE OF CONTENTS

1. SCOPE .....	4
2. APPLICABLE DOCUMENTS.....	4
2.1 SAE Publications .....	4
2.2 AIA Publications.....	4
3. OUTLINE OF METHOD.....	5
4. MATERIALS.....	5
5. APPARATUS.....	5
5.1 Filtration Apparatus.....	5
5.2 Particle Count Apparatus.....	6
6. REAGENTS .....	6
6.1 Aqueous.....	6
6.2 Alcohols .....	6
6.3 Rinse Liquid .....	7
6.4 Cleanliness .....	7
7. PREPARATION OF APPARATUS .....	7
8 LIQUID SAMPLES.....	7
9. FILTRATION PROCEDURE .....	8
9.1 Blank Analysis Filtration.....	8
9.2 Sample Filtration .....	8
10. MICROSCOPE CALIBRATION.....	9
11. PARTICLE COUNTING PROCEDURE.....	11
12. PARTICLE COUNT CALCULATION.....	13
13. COUNTING PROFICIENCY .....	13
14. DISPUTE CLAUSE .....	13