



Standard Test Method for Determination of Trace Thiophene in Refined Benzene by Gas Chromatography¹

This standard is issued under the fixed designation D4735; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the determination of thiophene in refined benzene in the range from 0.80 to 1.80 mg/kg for the Flame Photometric Detector (FPD), and from 0.14 to 2.61 mg/kg for the Pulsed Flame Photometric Detector (PFPD). For the PFPD, the minimum level of quantitation (LOQ) is 0.14 mg/kg and the minimum level of detection (LOD) is 0.04 mg/kg, as described in ASTM Research Report RR:D16-1038.² The range of the test method may be extended by modifying the sample injection volume, split ratios, calibration range, or sample dilution with thiophene-free solvent.

1.2 This test method has been found applicable to benzene characteristic of the type described in Specifications [D2359](#) and [D4734](#) and may be applicable to other types or grades of benzene only after the user has demonstrated that the procedure can completely resolve thiophene from the other organic contaminants contained in the sample.

1.3 The following applies to all specified limits in this test method: for purposes of determining conformance to applicable specification using this test method, an observed value or a calculated value shall be rounded off “to the nearest unit” in the last right-hand digit used in expressing the specification limit in accordance with the rounding-off method of Practice [E29](#).

1.4 The values stated in SI units are to be regarded as standard.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applica-*

bility of regulatory limitations prior to use. For specific hazard statements, see Section 7.

2. Referenced Documents

2.1 *ASTM Standards:*³

[D1193](#) Specification for Reagent Water

[D1685](#) Test Method for Traces of Thiophene in Benzene by Spectrophotometry (Withdrawn 2009)⁴

[D2359](#) Specification for Refined Benzene-535

[D3437](#) Practice for Sampling and Handling Liquid Cyclic Products

[D4057](#) Practice for Manual Sampling of Petroleum and Petroleum Products

[D4177](#) Practice for Automatic Sampling of Petroleum and Petroleum Products

[D4307](#) Practice for Preparation of Liquid Blends for Use as Analytical Standards

[D4734](#) Specification for Refined Benzene-545

[D6809](#) Guide for Quality Control and Quality Assurance Procedures for Aromatic Hydrocarbons and Related Materials

[E29](#) Practice for Using Significant Digits in Test Data to Determine Conformance with Specifications

[E177](#) Practice for Use of the Terms Precision and Bias in ASTM Test Methods

[E260](#) Practice for Packed Column Gas Chromatography

[E355](#) Practice for Gas Chromatography Terms and Relationships

[E691](#) Practice for Conducting an Interlaboratory Study to Determine the Precision of a Test Method

[E840](#) Practice for Using Flame Photometric Detectors in Gas Chromatography

[E1510](#) Practice for Installing Fused Silica Open Tubular Capillary Columns in Gas Chromatographs

¹ This test method is under the jurisdiction of ASTM Committee [D16](#) on Aromatic Hydrocarbons and Related Chemicals and is the direct responsibility of Subcommittee [D16.04](#) on Instrumental Analysis.

Current edition approved Feb. 1, 2014. Published February 2014. Originally approved in 1987. Last previous edition approved in 2009 as D4735 – 09. DOI: 10.1520/D4735-09R14.

² Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR:D16-1038. Contact ASTM Customer Service at service@astm.org.

³ For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

⁴ The last approved version of this historical standard is referenced on www.astm.org.

*A Summary of Changes section appears at the end of this standard

TABLE 1 Thiophene in Benzene Instrumental Conditions

Column	A	B	C	D
Tubing	6 ft 3 $\frac{1}{8}$ in. Ni	15 ft by $\frac{1}{8}$ in. stainless steel	10 ft by $\frac{1}{8}$ in. stainless steel	30 meter, Fused Silica, 0.25 (or 0.32) mm ID
Phase	TCEEP ^A	SP-1000	OV-351	Bonded Polyethylene Glycol (PEG/CW)
Concentration, weight %	7	10	10	0.5 (or 1.0) micron film thickness
Support	Chromosorb P-AW ^B	Supelcoport	Chromosorb P-AW	N/A
Mesh	100/120	60/80	80/100	N/A
Gas chromatographic conditions				
Inlet	150	170	180	200
Carrier Gas	helium	helium	helium	helium
Carrier Flow, mL/min	30	30	30	1.0–1.5
Split Ratio	N/A	N/A	N/A	50:1
Column Temperature, °C	70	90	70	50°C for 1 mi., 10°C/min to 200°C, hold for 1 min
Detector (optimize flows per manufacturer's instructions)	FPD	FPD	FPD	PFPD (tuned for Sulfur) BG-12 Filter 2 mm combustor
H ₂ , mL/min	140	140	140	11.5 flow optimized for S mode
Air 1, mL/min	80	80	80	12.0 flow optimized for S mode
Air 2, mL/min	70	70	70	10.0 flow optimized for S mode
Temperature (°C)	220	220	250	250

^A Tetracyanoethylated pentaerythritol or pentrile.

^B Chromosorb P is a registered trademark of the Manville Corp.

2.2 Other Document:

OSHA Regulations, 29 CFR paragraphs 1910.1000 and 1910.1200⁵

3. Summary of Test Method

3.1 The thiophene concentration in refined benzene is determined at the milligram thiophene per kilogram sample level using conventional gas-liquid chromatography with a flame photometric detector (FPD) or pulsed flame photometric detector (PFPD). A reproducible volume of sample is injected. Quantitative results are obtained by the external standard technique using the measured peak area of thiophene.

4. Significance and Use

4.1 This test method is suitable for setting specifications on benzene and for use as an internal quality control tool where benzene is either produced or used in a manufacturing process.

4.2 This test method was found applicable for determining thiophene in refined benzene conforming to the specifications described in Specification D2359 and may be applicable toward other grades of benzene if the user has taken the necessary precautions as described in the text.

4.3 This test method was developed as an alternative technique to Test Method D1685.

5. Apparatus

5.1 *Gas Chromatograph*—Any chromatograph having a flame photometric detector (FPD or PFPD) may be used which

can operate at the typical conditions described in Table 1. The user is referred to Practices E260 and E355 for additional information about gas chromatography principles and procedures. An automatic sampler is recommended. The GC should have the following performance characteristics:

5.1.1 *Column Temperature Programmer*—The chromatograph shall be capable of linear programmed temperature operation over a range sufficient for the separation of the compounds of interest. The programming shall be sufficiently reproducible to obtain retention time repeatability throughout the scope of the analysis.

5.1.2 *Sample Inlet System*—The sample inlet system shall have variable temperature control capable of operating continuously at a temperature up to the maximum column temperature employed. The sample inlet system shall allow a constant volume of sample to be injected by means of a syringe. For the PFPD a heated flash vaporizing injector designed to provide a linear sample split injection (that is, 50:1) is required for proper sample introduction. The associated carrier gas flow controls shall be of sufficient precision to provide reproducible column flows and split ratios in order to maintain analytical integrity.

5.2 *Column*—The column shall provide complete resolution of thiophene from benzene and any other hydrocarbon impurities because of potential quenching effects by hydrocarbons on the light emissions from the thiophene. The columns described in Table 1 have been judged satisfactory. The user is referred to Practice E1510 for assistance on installing fused silica capillary columns into the gas chromatograph.

5.3 *Detector*—Any flame photometric detector (FPD or PFPD) can be used, provided it has sufficient sensitivity to produce a minimum peak height twice that of the base noise for

⁵ Available from U.S. Government Printing Office Superintendent of Documents, 732 N. Capitol St., NW, Mail Stop: SDE, Washington, DC 20401, <http://www.access.gpo.gov>.