

DIN SPEC 91127

ICS 17.200.10

**Recommendation for Temperature Calibration of Fast Scanning Calorimeters (FSCs) for Sample Mass and Scan Rate;
Text in English**

Empfehlung für die Temperaturkalibrierung von Fast Scanning Calorimeters (FSCs)
für Probenmasse und Scanrate;
Text in Englisch

Recommandations pour l'étalonnage de la température des Calorimètres à Balayage
Rapide de la masse de l'échantillon et la vitesse de balayage;
Texte en anglais

Zur Erstellung einer DIN SPEC können verschiedene Verfahrensweisen herangezogen werden:
Das vorliegende Dokument wurde nach den Verfahrensregeln einer PAS erstellt.

Gesamtumfang 37 Seiten

Bereich Innovation

Contents

	Page
Foreword	3
Introduction	4
1 Scope.....	4
2 References	4
3 Abbreviations and Units	5
3.1 Abbreviations	5
3.2 Units	5
3.3 Convention.....	6
4 Fast Scanning Calorimetry (FSC)	6
4.1 General	6
4.2 DSCs.....	7
4.3 Chip Calorimeters.....	7
5 Temperature Calibration Protocol	8
5.1 FSC Performance (Annex A)	8
5.2 Primary Temperature Calibration (Annex B).....	8
5.3 Check of the Primary Temperature Calibration	9
5.4 Temperature Calibration with Respect to Sample Mass and Scan Rate (Annex C)	9
5.5 Check of the Temperature Calibration for Sample Mass and Scan Rate (Annex C)	10
5.6 Symmetry (Annex D)	10
6 Calibration Protocol Flow Chart.....	12
Annex A (informative) Examples of Performance.....	13
A.1 HyperDSCs PerkinElmer.....	13
A.2 RHC TA Instruments	16
A.3 FSC Universität Rostock.....	17
A.4 Flash DSC 1 Mettler-Toledo.....	18
Annex B (informative) Primary Temperature Calibration	19
B.1 HyperDSCs PerkinElmer.....	19
B.2 RHC TA Instruments	20
Annex C (informative) Temperature Calibration with Respect to Sample Mass and Scan Rate.....	21
C.0 Suitable sample masses for various FSCs	21
C.1 HyperDSCs PerkinElmer.....	22
C.2 RHC TA Instruments	25
C.3 FSC Universität Rostock.....	27
C.4 Flash DSC 1 Mettler-Toledo.....	28
Annex D (informative) Symmetry.....	29
D.1 HyperDSCs PerkinElmer.....	29
D.2 RHC TA Instruments	32
D.3 FSC Universität Rostock.....	33
D.4 Flash DSC 1 Mettler-Toledo.....	34
Bibliography	36

Foreword

This DIN Specification on differential scanning calorimetry, which has been developed by a standardization group using the PAS procedure, deals with high to very high scan rates and low to very low sample masses respectively, and is complementary to the International Standard ISO 11357 which is dealing with low scan rates.

It was elaborated by the following authors, members of the Standardization group¹⁾

- Dr Geert Vanden Poel, DSM Resolve, Geleen, The Netherlands;
- Dr Albert Sargsyan, DSM Resolve, Geleen, The Netherlands;
- Prof. Vincent Mathot, SciTe, the Netherlands and Katholieke Universiteit Leuven, Belgium, Chairman;
- Prof. Guy Van Assche, Vrije Universiteit Brussel, Belgium;
- Dr Andreas Wurm, Universität Rostock, Germany;
- Prof. Christoph Schick, Universität Rostock, Germany;
- Prof. Andres Krumme, Tallinn University of Technology, Estonia
- Prof. Dongshan Zhou, Nanjing University, People's Republic of China.

The following persons also contributed to the document:

- Dr Sander van Herwaarden, Xensor Integration, Delfgauw, The Netherlands;
- Elina Iervolino, Xensor Integration, Delfgauw, The Netherlands;
- Gunnar Schulz, Universität Rostock, Germany;
- Davit Zohrabyan, Universität Rostock, Germany;
- Evgeny Zhuravlev, Universität Rostock, Germany.

The work is an outcome of the NaPolyNet project of the 7th Framework Programme of the European Commission and was partially financed by NaPolyNet funds (EU-FP7 CSA NMP-2007-2.1-3).

DIN Deutsches Institut für Normung e.V., Berlin
March 21st, 2011

1) Although the authors strived to ensure that the information in this report is up-to-date and accurate as possible, they do not accept liability with regard to problems incurred as a result of using this report and/or any linked reports. The authors cannot be held liable for faults due to incorrectness, incompleteness and/or inaccuracy of the information in this report. In no event will the authors be liable for any damages whatsoever directly or indirectly arising out of or in connection with the use of the information in this report. The authors may revise the information in this report without notice. This disclaimer and report are construed under and shall be governed by the laws of Germany.