



# Standard Test Method for Purity of 1,3-Propanediol (Gas Chromatographic Method)<sup>1</sup>

This standard is issued under the fixed designation D 7515; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This test method describes the gas chromatographic determination of purity for 1,3-propanediol (PDO). This test method was originally developed to determine the purity of 1,3-propanediol used for the application as the freeze point depressant base fluid in formulated PDO engine coolants. Use of the method for purity of PDO for other applications may be viable.

1.2 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.3 Review the current Material Safety Data Sheets (MSDS) for detailed information concerning toxicity, first aid procedures, and safety precautions.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

## 2. Referenced Documents

2.1 *ASTM Standards:*<sup>2</sup>

**E 177 Practice for Use of the Terms Precision and Bias in ASTM Test Methods**

**E 300 Practice for Sampling Industrial Chemicals**

**E 691 Practice for Conducting an Interlaboratory Study to Determine the Precision of a Test Method**

## 3. Summary of Test Method

3.1 The neat sample is analyzed by a temperature-programmed gas chromatograph, equipped with a capillary

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee D15 on Engine Coolants and is the direct responsibility of Subcommittee D15.07 on Specifications. Current edition approved April 1, 2009. Published May 2009.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

column and flame ionization detector (FID), and quantification is performed by direct area normalization.

3.2 Additionally, the use of a reference sample using Ethylene, Propylene or Dipropylene Glycol (EG, PG or DPG) in 1,3-PDO (minimum purity 99.5 %) should be used as a performance check (see Section 8).

NOTE 1—The application of this reference sample is also used to demonstrate the separation of commonly used glycols (EG, PG and DPG) in engine coolants, from PDO. Solutions of EG, PG, or DPG in concentrations of 0.1 to not more than 1 % may be used.

## 4. Significance and Use

4.1 Knowledge of an approved method is required to establish whether the product meets the requirements of its specifications. The use of glycols in the reference sample is not intended to suggest the presence of glycol (EG, PG and DPG) impurities, but to demonstrate and quantify the separation of commonly used Engine Coolant glycols from PDO.

## 5. Apparatus

5.1 *Gas Chromatograph(s)*—provided with a sample splitter or on-column injection, flame ionization detector and temperature-programming facilities. The instrument must be suitable for analysis according to the operating instructions given in Table 1. To account for differences among laboratory equipment, the two most common column choices are listed.

NOTE 2—Other column suppliers market alternative stationary phases, therefore, it is permissible to use a different column from an alternative supplier. However, the chromatogram obtained must be identical, with regard to separation of PDO and other glycol components, to those illustrated in Fig. A1.1 and Fig. A1.2.

5.1.1 *Columns*—The analytical column used must completely separate EG, PG or DPG from PDO. Fig. A1.1 and Fig. A1.2 show examples of chromatograms conforming to the requirements.

5.2 *Digital Integration Equipment*—A computer with data collection software.

5.3 *Analytical Balance*, readability 0.1 mg, calibrated. Calibrate and verify at regular intervals.

**TABLE 1 Typical Operating Parameters for the GC Analysis of PDO**

| Column <sup>A</sup> | Option A   | Option B   |
|---------------------|--|--|
| Type                | Capillary  | Capillary  |
| Material            | Fused Silica   | PEG  |
| Length × I.D.       | 10 m × 0.1 mm  | 30 m × 0.25 mm   |
| Stationary Phase    | DB-5   | ZB-Wax   |
| Film Thickness      | 0.17 μm  | 0.25 μm  |
| Detector System     |  |  |
| Type                | FID  | FID  |
| Sensitivity         | The ratio of the signal to the noise level must be at least 2:1 at a concentration of 5 mg/kg glycols in PDO | The ratio of the signal to the noise level must be at least 2:1 at a concentration of 5 mg/kg glycols in PDO                               |
| Temperatures        |  |  |
| Column Oven         |  |  |
| Initial             | 0.5 min at 35°C  | 0 min at 50°C  |
| Ramp 1              | 35 to 85°C at 50°C/min   | 50 to 200°C at 15°C/min  |
| Ramp 2              | 85 to 325°C at 100°C/min   | 200 to 250°C at 40°C/min   |
| Ramp 3              | 2 min at 325°C   | 17 min at 250°C  |
| Detector            | 325°C  | 250°C  |
| Carrier Gas         | Helium   | Helium   |
| Calibration         | This method employs straight area normalization so no calibration is required                                | This method employs straight area normalization so no calibration is required  |
| Injected Volume     | 0.1 μL   | 0.2 μL   |
| Pressure Program    | 0.5 min 30 psi<br>30 to 100 psi at 100 psi/min<br>8 min at 100 psi<br>Gas saver on at 0.5 min                | Pressure: 13.2 psi at 50°C<br>Flow: 1.1 mL/min<br>Velocity: 28 cm/s  |
| Split Ratio         | 1:250 or appropriate split ratio to allow adequate sensitivity as defined under Detector System              | 1:18 or appropriate split ratio to allow adequate sensitivity as defined under Detector System (only if split injection technique is used) |

<sup>A</sup> The columns are available commercially. Some column suppliers market alternative stationary phases. The chromatogram obtained must be identical, with regard to separation of PDO and other glycol components, to those illustrated in Fig. A1.1 and Fig. A1.2.

5.4 *Crimp Top Vials*, 1 mL and 5 mL.

5.5 *Crimper/De-capper*, for capping and de-capping the vials.

5.6 *Micro Syringes*, 5 μL or 10 μL.

5.7 *Bottles*, 100 mL, with screw cap.

## 6. Reagents and Materials

6.1 *Purity of Reagents*—Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society where such specifications are available.<sup>3</sup> Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

### 6.2 Reagents:

6.2.1 *1,3-Propanediol (PDO)*, minimum purity 99.5 % mass (m/m).

6.2.2 *Ethylene Glycol (EG)*, minimum purity 99.5 % mass (m/m).

6.2.3 *Propylene Glycol (PG)*, minimum purity 99.5 % mass (m/m).

<sup>3</sup> *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeia Convention, Inc. (USPC), Rockville, MD.

6.2.4 *Dipropylene Glycol (DPG)*, minimum purity 99.0 % mass (m/m).

6.3 *Water*, HPLC grade.

## 7. Sampling, Test Specimens and Test Units

7.1 Follow the relevant instructions for sampling as given in Practice E 300.

## 8. Preparation of Apparatus

8.1 *Gas Chromatograph(s) and Column(s)*—Check the performance of the gas chromatograph and column as follows:

8.2 Using the standard quality reagents (6.2), prepare a 1,3-PDO solution containing approximately 0.1 % of EG, PG and DPG respectively. Determine the exact concentration of the components. This will be the reference sample.

8.2.1 Weigh 0.1 g of each glycol reagent to the nearest 0.1 mg, into a 100-mL vial. Add 99.7 g of 1,3-PDO weighed to the nearest 0.1 mg. Cap the vials and mix thoroughly.

8.2.2 Calculate the exact concentration of each glycol in the reference sample.

8.3 Fill a 1-mL GC autosampler vial with the reference sample (8.2) and close the vial.

8.4 Analyze the reference sample using the parameters given in Table 1. Inject the solution at least twice. Calculate the area %.