

# INTERNATIONAL STANDARD ISO 13366-1 IDF 148-1:2008 TECHNICAL CORRIGENDUM 1

Published 2009-08-01

INTERNATIONAL ORGANIZATION FOR STANDARDIZATION • MEЖДУНАРОДНАЯ ОРГАНИЗАЦИЯ ПО СТАНДАРТИЗАЦИИ • ORGANISATION INTERNATIONALE DE NORMALISATION

## Milk — Enumeration of somatic cells —

## Part 1:

# Microscopic method (Reference method)

**TECHNICAL CORRIGENDUM 1** 

Lait — Dénombrement des cellules somatiques —

Partie 1: Méthode au microscope (Méthode de référence)

RECTIFICATIF TECHNIQUE 1

Technical Corrigendum 1 to ISO 13366-1 IDF 148-1:2008 was prepared by Technical Committee ISO/TC 34, Food products, Subcommittee SC 5, Milk and milk products, and the International Dairy Federation (IDF). It is being published jointly by ISO and IDF.

Page 5, 7.2

In the denominator on the right-hand side of the equation, delete "x", insert "+", so that it reads:

$$d = \frac{V_{s}}{V_{s} + V_{b}}$$

ICS 67.100.10

Ref. No. ISO 13366-1:2008/Cor.1:2009(E)

IDF 148-1:2008/Cor.1:2009(E)

#### Page 11, 9.2, Equation (2) and its variable definitions

In the numerators of the right-hand side of Equation (2) and in the second equation after it, delete " $W_s$ ", insert " $L_s$ ", so that the three equations read:

$$c = \frac{L_{\rm S} N_{\rm t}}{D_{\rm f} N_{\rm b} V_{\rm m}} \cdot \frac{1}{d} \tag{2}$$

or

$$c = f_{\mathbf{W}} \cdot \left( \frac{N_{\mathsf{t}}}{N_{\mathsf{b}}} \cdot \frac{\mathsf{1}}{d} \right)$$

with the constant working factor,  $f_{w}$ 

$$f_{\rm W} = \frac{L_{\rm S}}{D_{\rm f} V_{\rm m}}$$

### Page 14, Table A.1

In column 1, row 3 (excluding the header row), second line, delete "Coefficient of variation of repeatability (%)" and insert "Coefficient of variation of standard deviation of repeatability (%)".

In column 1, row 4 (excluding the header row), second line, delete "Coefficient of variation of reproducibility (%)" and insert "Coefficient of variation of standard deviation of reproducibility (%)".

In column 2, row 3 (excluding the header row), second line, delete "169", insert "16".

Table A.1 as corrected now reads:

Table A.1 — Results of interlaboratory test

	Level			
	1	2	3	4
No. of participants after eliminating outliers	24	23	24	24
Mean value (x 1 000 cells/ml)	245	455	679	791
Repeatability standard deviation, $s_r$ (× 1 000 cells/ml)	38	43	69	110
Coefficient of variation of standard deviation of repeatability (%)	16	9	10	14
Repeatability limit, $r$ (2,8 $s_r$ ) (× 1 000 cells/ml)	107	121	192	308
Reproducibility standard deviation, $s_R \times 1000 \text{ cells/ml}$	41	62	78	110
Coefficient of variation of standard deviation of reproducibility (%)	17	14	11	14
Reproducibility limit, $R$ (2,8 $s_R$ ) (× 1 000 cells/ml)	114	174	218	308