

INTERNATIONAL STANDARD

ISO 9622

First edition
1999-11-15

Whole milk — Determination of milkfat, protein and lactose content — Guidance on the operation of mid-infrared instruments

*Lait entier — Détermination des teneurs en matière grasse laitière,
en protéines et en lactose — Lignes directrices pour l'utilisation
des appareils de dosage par absorption dans le moyen infrarouge*



Reference number
ISO 9622:1999(E)

Contents

1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 Principle	2
5 Principal characteristics of infrared instruments	2
6 Factors affecting the accuracy of measurement	3
6.1 Instrument factors	3
6.2 Physicochemical and biological factors	5
7 Calibration of the instrument	7
7.1 Objective	7
7.2 Checking the initial calibration for fat, protein and lactose	7
7.3 Maintaining calibration and confirmation of calibration validity	8
8 Sampling	8
9 Uniformity of test samples	8
10 Determination	8
11 Checking daily short-term stability of the instrument	9
11.1 General	9
11.2 Preparation and storage of sub-samples	9
11.3 Analysis of control samples	9
11.4 Monitoring the analytical procedure	9
11.5 Re-adjustment of calibration	11
12 Precision and accuracy	11
12.1 Repeatability	11

© ISO 1999

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

International Organization for Standardization
Case postale 56 • CH-1211 Genève 20 • Switzerland
Internet iso@iso.ch

Printed in Switzerland

12.2 Intralaboratory reproducibility 12

12.3 Accuracy..... 12

13 Test report 12

Annex A (informative) Procedure for setting and evaluating linearity on a mass/volume basis 13

Annex B (informative) Control and adjustment of correction factors 15

Annex C (informative) Procedure for calibration of infrared milk analysers using modified milk samples 24

Bibliography..... 27