
**Water quality — Determination of tetra- to
octa-chlorinated dioxins and furans —
Method using isotope dilution
HRGC/HRMS**

*Qualité de l'eau — Dosage des dioxines et furanes tétra- à
octachlorés — Méthode par dilution d'isotopes HRGC/SMHR*



Reference number
ISO 18073:2004(E)

© ISO 2004

PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

© ISO 2004

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword	v
1 Scope	1
2 Normative references	1
3 Terms, definitions and abbreviated terms	2
3.1 Terms and definitions	2
3.2 Abbreviated terms	3
4 Principle	4
4.1 Spiking and extraction	4
4.2 Clean-up	4
4.3 Concentration	4
4.4 Identification	5
4.5 Quantification	6
4.6 Analytical quality	6
5 Contamination and interferences	6
6 Reagents and standards	8
7 Apparatus and materials	15
7.1 Water sample containers	15
7.2 Equipment for sample preparation	15
7.3 Extraction apparatus	15
7.4 Filtration apparatus	15
7.5 Clean-up apparatus	16
7.6 Concentration apparatus	16
7.7 Other equipment	17
8 Sample collection, preservation, storage and holding times	17
9 Quality assurance/quality control	18
9.1 General	18
9.2 Initial precision and recovery (IPR)	20
9.3 Spiking	20
9.4 Method blanks	21
9.5 Quality control check sample	21
9.6 Method precision	21
10 Calibration	22
10.1 Operating conditions	22
10.2 Mass spectrometer (MS) resolution	23
10.3 QC limits and theoretical ion abundance ratios for standards	25
10.4 Retention time	25
10.5 Isomer specificity	26
10.6 Calibration by isotope dilution	27
10.7 Calibration by internal standard	28
10.8 Combined calibration	29
11 Sample preparation	29
11.1 General	29
11.2 Determination of suspended solids	29
11.3 Preparation of aqueous samples containing 1 % or less by mass suspended solids	30
12 Extraction and concentration	31
12.1 Extraction of samples with no visible particulates	31
12.2 Soxhlet extraction of filters and/or disks	31